

Shaker incubator/growth chamber for algae

Date: 12-09-2019

This is a clarification note for the above-mentioned tender placed on IISc website on 05-09-2019.

Addendum to technical specifications of the web tender:

10. Chamber Dimensions: 60 x 62 x 170 cm or larger
11. Platform Size of the shaker: 46 × 46 cm or larger
12. Shaker agitation orbit: 1.9 cm or larger
13. Shaker should be able to continuously shake for 96 hours or more.

Terms and conditions of the tender:

1. The technical specifications mentioned in the tender are highly desired. However, lower technical specifications may be considered if the above mentioned specifications are found to be unsuitable in financial terms. The Institute reserves the right to go for lower specifications taking into consideration its financial constraints and technical preferences.
2. Please enclose a compliance certificate of the product that is quoted.
3. The vendor should have a good track record of having previously supplied similar equipment in India (please furnish the details).
4. The vendor should have qualified technical service personnel for the equipment based in India (preferably in Bangalore).
5. The lead time for the delivery of the equipment should not be more than 3 months from the date of receipt of purchase order.