



**Office of Deputy Registrar (Purchase)**  
**Indian Institute of Science (IISc)**  
**Bangalore – 560012**  
(<https://www.iisc.ac.in>)

**CORRIGENDUM-1**

No.: IISc/Purchase/CDC/2020/14/Corr-1 Date: 5<sup>th</sup> August 2020

Ref. Tender No: IISc/Purchase/CDC/2020/14, Dated 27<sup>th</sup> July 2020

CPPP Tender ID: 2020\_IISC\_573901\_1

**NOTICE INVITING TENDER**

in e-Tender mode only on the  
Central Public Procurement Portal (CPPP) of the Govt. of India  
under the two-cover bid system

*for*

**Supply and Installation of**  
**Universal 3-D Coordinate Measuring Machine (CMM) at**  
**IISc - Challakere Campus**

Tender No: IISc/Purchase/CDC/2020/14

Date: 27<sup>th</sup> July 2020

**Contact Details for this tender:**

Chair

Challakere Empowered Committee  
Challakere Development Centre Office  
Indian Institute of Science  
Bangalore 560012

**Email: [bgm@iisc.ac.in](mailto:bgm@iisc.ac.in)**

**Online Tender Submission website:**

<https://eprocure.gov.in/eprocure/app>

**CORRIGENDUM-1****Tender Ref. No.: IISc/Purchase/CDC/2020/14, Dated 27<sup>th</sup> July 2020**

Indian Institute of Science, Bangalore has issued a Notice Inviting Tender (Tender no. IISc/Purchase/CDC/2020/14, Dated 27<sup>th</sup> July 2020, CPPP Tender ID: 2020\_IISC\_573901\_1) for Supply and Installation Universal 3-D Coordinate Measuring Machine (CMM )at IISc - Challakere Campus, Indian Institute of Science. In this tender, based on the pre-bid queries, the Purchase Committee has decided certain amendment in the technical specifications: -

1. Following are changes in the existing content / technical specifications of the NIT: -

<b>Sl. No.</b>	<b>Reference in the NIT</b>	<b>Existing Content in the NIT</b>	<b>Revised / Amended Content (Now READ AS)</b>
1.	Page – 3, a) Primary Specifications – S. N. 7	<del>Drive acceleration: 2300 mm/sec<sup>2</sup> or better</del>	Drive acceleration: 1500 - 2300 mm/sec <sup>2</sup>
2.	Page – 4, a) Primary Specifications – S. N. 17	<del>Power supply: 220-230 Volts single phase, 50-60 Hz</del>	Power supply: Power supply with surge protection and stabilization
3.	Page – 4, a) Primary Specifications – S. N. 18	<del>Load carrying capacity of work table : 1500 kgs (min)</del>	Load carrying capacity of table : 1000 to 1500 kg

2. All other terms & conditions and rest of the contents / technical specifications of the Notice Inviting Tender (NIT) will remain unchanged. This corrigendum forms an integral part of the NIT. The bidders are required to submit a copy of the NIT and this corrigendum (duly sealed and signed of the bidder on all pages) as an annexure with their Technical Bid. All bids must be submitted through e-tender mode only via CPPP.
3. Bidders may regularly visit websites [www.iisc.ac.in](http://www.iisc.ac.in) and <https://eprocure.gov.in/eprocure/app> for any further details with regard to this tender.
4. All query / correspondence regarding this tender should be addressed to “The Chair, Challakere Empowered Committee, Challakere Development Centre Office, Indian Institute of Science, Bangalore – 560012, India” (Email ID: [bgm@iisc.ac.in](mailto:bgm@iisc.ac.in)) only. Query / correspondence addressed to any other official / authority of IISc will not be entertained.

Chair  
Challakere Empowered Committee