

To Whom It May Concern

Domestic Tender for a High temperature 3 zone horizontal tube furnace

This is an RFQ (Request for Quote) for the procurement of a High temperature 3 zone horizontal tube furnace system as part of a domestic tender for the Centre for Nano Science and Engineering (CeNSE) at IISc, Bangalore. The tender invitation is for Indian Original Equipment Manufacturer (OEM)/Class-1/Class-2 or their Indian authorized distributor only.

CeNSE is a multidisciplinary research department at IISc that houses a 14,000 sq. ft. cleanroom and characterization facility used by 50 faculty members from various disciplines at IISc. Consequently, any tool in CeNSE receives significant exposure to the scientific community at IISc and beyond. The vendors are requested to factor in the value of this exposure into their quotes.

http://nnfc.cense.iisc.ac.in/ http://www.mncf.cense.iisc.ac.in/ https://www.inup.cense.iisc.ac.in/

Procedure

- 1. Vendors will be required to submit a technical proposal and a commercial proposal in **two separate sealed envelopes**. Only vendors who meet the technical requirement will be considered for the commercial negotiation.
- 2. The deadline for submission of proposals is the 8th July 2022, 5:30 pm Indian Standard Time. Proposals should arrive at the Centre for Nano Science and Engineering (CeNSE), Indian Institute of Science, Bangalore 560012, India, by the above deadline.
- 3. The decision of the purchase committee will be final.
- 4. The quotations should be on FOR-IISc Bangalore basis in INR only.
- 5. The commercial comparison will be made as per Government of India rules, specifically GFR 2017. Note that GFR has recently been amended.
- 6. As per recent edits to the GFR, there are three classes of vendors distinguished by their "local content". In the cover letter, vendors must mention which applies to them: Class 1 supplier: Goods and services have a local content of equal to or more than 50% Class 2 supplier: Goods and services have a local content of more than 20% but less than 50% Non-local supplier: Goods and services have a local content of equal to or less than 20% 5. Quotes will be entertained from Class 1 or Class 2 suppliers only.
- 7. The technical proposal should contain a compliance table with 5 columns. The first column must list the technical requirements in the order that they are given in the technical configuration below. The second column should describe your compliance in a "Yes" or "No" response. If "No", the third column should provide the extent of the deviation (please provide quantitative responses). The fourth column should state the reasons for the deviation if any. The fourth column can be used to compare your tool with that of your competitors or provide details as requested in the technical requirements table below.



- 8. Any additional capabilities or technical details that you would like to bring to the attention of the purchase committee can be listed at the end of the technical table.
- 9. Vendors are encouraged to highlight the advantages of their tools over comparable tools from the competitors.
- 10. If multiple systems can fulfil the requirements, vendors can submit multiple bids.
- 11. In the commercial bid, please provide an itemized cost of the system and required accessories, such as software, power supply, etc.
- 12. As an option, please provide itemized cost for any suggested accessories/add-ons that may enhance the usability, capability, accuracy or reliability of the tool. Vendors are encouraged to quote for as many add-ons as their tool portfolio permits.
- 13. The quotes should include all charges for door delivery.
- 14. Please indicate the warranty provided with the tool. A warrant of 2 years is preferred.
- 15. Provide itemized cost for required spares for 2 years of operation. For the sake of this calculation, the vendor may assume active tool usage of 20 hours/ week. This number will be used to estimate the life cycle cost of the tool.
- 16. If maintenance requires, as an additional option, provide a cost of an annual maintenance contract (AMC) for 3 years, post-warranty. The **AMC must cover 2 scheduled and 2 emergency visits per year**. The AMC cost must also include an itemized list of spares that are essential for the scheduled visits.
- 17. The **RFQ** must include references of a minimum of 5 previous installations, preferable in India. Please provide the names and contact addresses of the referees so that the committee can contact them independently.
- 18. Any technical questions can be directed to Prof. Pavan Nukala, CeNSE, IISc., Bengaluru (pnukala@iisc.ac.in) and Prof. Srinivasan Raghavan, CeNSE, IISc., Bengaluru (sraghavan@iisc.ac.in).

	Technical Requirement I. Annealing process using 1800 °C single zone horizontal tube furnace.					
1.	Temperature	• Maximum continuous operation temperature: 1700 °C and Maximum operating temperature: 1800 °C.				
2.	Furnace construction /design	 High temperature tube furnace designed for horizontal mode operation. Rectangular housing with holes for convection cooling. Furnace with separate control box with 3 m cable, plug and socket. 				
3.	Heating elements	• MoSi ₂ heating elements installed in rectangular housing and hanging vertically that can be easily replaced.				
4.	Heating zones	Single Zone.				
5.	Heating and Colling rates	 Heating: Minimum 0.5 °C and Maximum 5 °C Cooling: Gradual cooling down 				



		<u> </u>
6.	Insulation	 Low thermal mass ceramic fibre insulation. High grade insulation material with low thermal conductivity consisting of ceramic fibre inner insulation, ceramic fibre case insulation and ceramic fibre end insulation. Installed in a rectangular housing to provide low energy consumption and high heating rates.
7.	Configuration	Horizontal mode
8.	Thermocouple	High grade type B thermocouple.
9.	Over Temperature Protection Control	Over Temperature Protection Control to protect load or furnace during unattended operation.
10.	Programmable temperature controller	 Colour Touch screen controller offering Setpoint control, Program profile control, 10 unique program profiles saved in memory, 24 segments per unique program, ethernet communication, Panel mounted USB socket, Data logging to a USB memory stick in a .csv file format, Real time clock, Program schedule start, Program status indication with estimate end time & date, 2 Events indication, Control power indication, User level security, Single point temperature calibration that can be set as ramp, step or dwell and can be configured to control relays.
11.	Heated length	 Heated tube maximum length upto (mm): 300. Uniform length ±5°C (mm): 150.
12.	Dimensions	 Furnace external dimensions: H x W x D (mm): 600 ± 20 x 600 ± 20 x 500 ± 20 and weight ≤ 65 kg Control module dimensions H x W x D in mm: 850 ± 10 x 550 ± 10 x 500 ± 10 and weight ≤ 150 kg.
13.	Work tube package for operation of the furnace in inert atmosphere	 Recrystallized Alumina (C799) material work tube with 70 mm outer diameter, 60 mm inner diameter & 1115 mm length. Radiation shields – 2 Nos. A set of two insulation plugs.



		• Insulation sleeves – 2 Nos.		
		Water cooled high vacuum flanges		
		Furnace mounted flange holders		
		Inert gas inlet valve		
		• Flowmeter (argon 2 -20 litres per hour) – 1 No.		
		Over pressure valve.		
		• Insulation plugs – 2 Nos.		
14.	Power supply and maximum power (W)	• Three phase, 6400 W		
15.	Warranty	• 24 months from the date of commissioning and acceptance of equipment.		
II. Chemical Vapour Transport (CVT) process using 1200 Degrees C, Three zone Gradient horizontal split tube furnace				
16.	Temperature	• Maximum Temperature:1200°C. Maximum continuous operating temperature:1100 °C.		
17.	Heating elements	High quality heating elements with excellent and unsurpassed temperature uniformity along the entire heated length. Fast heat-up and cool-down rates.		
18.	Insulation	High quality thermal insulation		
19.	Heating zones	Three independently handling Zone.		
20.	Heated zones and unheated zones	Three 150 mm long heated zones and two 75 mm long unheated zone barriers.		
21.	Heating and Colling rates	Heating: Minimum 0.5 °C and Maximum 5 °C		
		Cooling: Gradual cooling down		
22.	Furnace type and dimensions	• Compact, horizontal and split type. Furnace body is split into two halves and hinged at the rear; pneumatic dampening struts at either end provide a smooth opening action.		
		• Furnace body dimensions: H x W x D in mm 575 ± 20 x 800 ± 20 x 500 ± 20. Control module dimensions: 225 ± 20 x 800 ± 20 x 500 ±20.		
\perp				



			XXX 1.1. 0.1. 0
		•	Weight of the furnace: Approximately 55 kg
23.	Furnace design and configuration	•	Flexible design to use a variety of tube diameters with the use of tube
			adapters. Specifically designed to provide a temperature gradient along the
			length of the three heated zones. Tube furnace mounted on top of the
			control box. The furnace body to be easily detached and separated for
			remote operation. Configuration to include a 2metre cable (including plug
			and socket) between the furnace body and control box.
24.	Temperature controllers and thermocouples.	•	Programmable temperature controllers to be fitted in three heated zones –
			3 Nos. One programmable temperature controller and one thermocouple
			to be fitted in each heated zone.
		•	Programmable temperature controller with 1 program and 24 segments.
			Each segment can be set as ramp, step or dwell and can be configured to
			control two relays.
		•	Ethernet communication is fitted as standard and certified for
			cybersecurity communications robustness.
25.	Temperature Gradients	•	Zone 1: 1100 °C, Zone 2: 950 °C, Zone 3: 800 °C.
		•	Zone 1: 1100 °C, Zone 2: OFF, Zone 3: 400 °C.
		•	Above temperature, gradients are to be achieved in 3 zone gradient tube
			furnace and graphical documentation confirming above data to be
			produced along with technical literature.
26.	Heating elements	•	Wire elements in high-quality vacuum-formed insulation ensure fast heat up, excellent temperature control and short cool down times.
			up, excellent temperature control and short cool down times.
27.	Over Temperature Protection Controller	•	Digital over-temperature protection controllers fitted in all three zones of
			the furnace. Includes three independent thermocouples fitted in all the three zones.
28.	1 0	Wo	ork tube package for the operation of the furnace in an inert atmosphere
	for operation in inert atmosphere	coı	nsisting of the following items to be included in the furnace.



		 Quartz work tube with 60 mm outer diameter, 55 mm inner diameter & 1050 mm length. Insulation plugs – 2 Nos. Work tube end seals: Gas inlet/outlet pipe – 1 No. Gas inlet/outlet pipe + thermocouple gland – 1 No. Probe thermocouple access up to 1200°C: Probe thermocouple gland complete with type N thermocouple.
29.	Standard inert gas package for operation in inert atmosphere.	 Gas inlet = 6 mm outside diameter push-in fitting Manual on/off ball valve Pressure relief valve Pressure gauge Flow meter with flow adjustment knob Non-return valve Fitting and pipe to connect an additional inert gas package. Gas outlet = 6 mm braided hose with 6 mm union
30.	Tube guards and tube supports	• Guards and horizontal tube support for extended length work tube for operation in inert gas operation – 1 set.
31.	Max power (W)	• Single Phase ≤ 2000 W
32.	Warranty	• 24 months from the date of commissioning and acceptance of equipment.
33.	Eligibility Criteria	The bidder must have supplied a minimum of three 1800 °C horizontal tube furnaces and three 1200 °C gradient tube (CVT) furnaces to IITs, IISERs, NITs, Government Institutes, and Reputed Process Industries/Research Institutes, preferably in India. Please provide the names and contact addresses of the referees so that the committee can contact them independently.



- Equipment offered must be a model from the current serial production range of the manufacturer. Customized or One-off Manufactured models will not be accepted. The offer should be supported with a printed catalogue & depiction on the company website. System Catalogue should be produced with the Technical Bid.
- PO copies to be submitted as proof.
- Manufacturers are requested to provide ISO certificate and Declaration on CE Conformity with minimum 3 valid directives. (The certificates need to be uploaded/submitted in the Statutory Cover.)
- CE Certification must be provided for the proposed system. The CE certificate should be provided with the Unit.
- Supplier should support the user with all the spares for a minimum period of 10 years.
- Details of an experienced service engineer, including contact detail, should be provided in the tender document. The supplier should mention the details of service setup and man powers who are responsible for aftersales support. Response time should be within 24 hrs.
- Up to date sales tax or GST clearance certificate.