

# **GLOBAL TENDER NOTICE**

## **Supply and Installation of High Shear Homogenizer**



**TENDER DOCUMENT**

**IISC/BE/RA/25-26/03**

**Department of Bioengineering  
Indian Institute of Science, Bangalore**

### Table of contents

|           |                      |                                                                 |
|-----------|----------------------|-----------------------------------------------------------------|
| Section 1 | Bid Schedule         |                                                                 |
| Section 2 | Eligibility Criteria | As specified by IISc                                            |
| Section 3 | Terms and Conditions | As specified by IISc                                            |
| Section 4 | Specifications       | Technical Specifications                                        |
| Section 5 | Technical Bid        | Annexure I                                                      |
| Section 6 | Commercial Bid       | Quotation with Price, Technical specifications of the equipment |

### BID SCHEDULE

|   |                                          |                                                                                                                                                                                                                         |
|---|------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 1 | Tender Number                            | IISC/BE/RA/25-26/03                                                                                                                                                                                                     |
| 2 | Tender date                              | 1 <sup>st</sup> August 2025                                                                                                                                                                                             |
| 3 | Item description                         | Supply and installation of High Shear Homogenizer                                                                                                                                                                       |
| 4 | Tender type                              | Two bid system:<br>a. Technical Bid (Part A)<br>b. Commercial Bid (Part B)                                                                                                                                              |
| 5 | Place of tender submission               | Office,<br>Department of Bioengineering<br>TCS-X Smart Hub<br>Indian Institute of Science,<br>Bangalore<br>560012<br>080-2293-2624                                                                                      |
| 6 | Last date and time for tender submission | 25 <sup>th</sup> August 2025, 5:00 PM                                                                                                                                                                                   |
| 7 | For further clarification                | Prof. Rachit Agarwal<br>Department of Bioengineering<br>TCS-X Smart Hub<br>Indian Institute of Science,<br>Bangalore 560012<br>Email: <a href="mailto:rachit@iisc.ac.in">rachit@iisc.ac.in</a><br>Phone: +91 7483315635 |

## **Eligibility Criteria**

### **Prequalification criteria:**

1. The Bidder's firm should have existed for at least 5 years. The bidder should enclose company registration certificate.
2. Only the Original Equipment Manufacturer or their authorized representatives across the globe shall participate in the bid.
3. The Bidder must not be blacklisted/banned/suspended or have a record of any service-related dispute with any organization in India or elsewhere. A declaration to this effect has to be given as per Annexure 3.
4. The bidder should sign and submit the declaration for Acceptance of Terms and Conditions as per Annexure 4.
5. The bidder should provide a support certification as per Annexure 6
6. The Bidder should have qualified technical service personnel for the instrument(s) based in India.
7. Foreign currency quotations should be on CIP Bangalore basis. INR quotes should be on a FOR-IISc Bangalore basis.
8. The order will be placed only on the bidder who participated in the bid.
9. The turnover for the last 2 years should be more than Rs. 10 crore. Bidder shall have to submit audited accounts (Balance sheet profit and loss account) of last two financial year. Audited statement must be signed and stamped by qualified chartered accounted.

## **Terms and conditions**

### **A. Submission of Tender:**

1. All documentations in the submission should be in English.
2. The bid should be submitted in the two-cover system, i.e., technical bid and commercial bid separately in sealed covers.
  - a. Technical Bid (Part-A) – The technical proposal should contain a compliance table besides the technical specifications listed in the description section below.

The compliance table should include all the items and in the same order. The first column should describe your compliance in a “Yes” or “No” response. The second column should state the extent of deviation if the answer was no or describe the exact specifications if the answer is yes. The “third” column should state the reasons for the deviation if the answer was No. The fourth column can be used to compare your solution with that of your competitors or provide details as requested in the technical requirements table below. Suppliers who include any indication of prices in the technical bid will be automatically disqualified.
  - b. Commercial Bid (Part-B) – Commercial bid indicating item wise price breakdown for the items mentioned in the technical bid, as per the format provided in tender, and other commercial terms and conditions. The price should be inclusive of all discounts and installation and training.
  - c. The technical bid and price bid should each be placed in separate sealed covers, superscripting on both the envelopes the tender no. and the due date. Both these sealed covers are to be placed in a bigger cover which should also be sealed and duly superscripted with the Tender No, Tender Description & Due Date.
  - d. The SEALED COVER superscripting tender number / due date should reach the address mentioned above on or before due date mentioned in the tender notice. In case the due date happens to be a holiday, the tender will be accepted and opened on the next working day. If the quotation cover is not sealed, it will be rejected.
3. GST/other taxes, levies etc., are to be indicated separately. The BIDDER should mention GST Registration and PAN in the tender document and provide a copy of both documents (Indian Bidders only).
4. If price is not quoted in Commercial Bid as per the format provided in tender document the bid is liable to be rejected.
5. The Institute reserves the right to accept or reject any bid and to annul the bidding process and reject all bids at any time prior to the award of contract, without there by incurring any liability to the affected bidder or bidders or any obligation to inform the affected bidder or bidders.

6. The technical proposal must include references of at least 10 previous installations done in India within last 5 years of similar equipment from the equipment manufacturer. Please provide the names and contact addresses of the five independent referees so that the committee can contact them independently to get a reference.
7. A technical evaluation by the purchase committee may include a demonstration to verify the functionalities and capabilities of the system quoted. The purchase committee reserves the right to reject the bids based on their technical evaluation of the quality of data, capability demonstration, and service. If the data/requested capability demonstration does not happen within a stipulated timeframe, the bid will be rejected. Any discrepancy between the promised specifications and measurements will be deemed as technical non-compliance. Imported items should be shipped on C.I.P. Bangalore basis (by Air Freight only), and all components and accessories indicate component-wise and itemized breakup. Provide certificates for the country of origin of manufacturing for each line item. The price of every line item in the commercial bid should be quoted along with the total quoted price for the instrument to be operational (installed and ready to use) in our facility.
8. The purchase committee reserves the right to accept or reject any bid and to annul the bidding process and reject all bids at any time prior to the award of contract, without there by incurring any liability to the affected bidder or bidders or any obligation to inform the affected bidder or bidders.
9. Incomplete bids will be summarily rejected.
10. The decision of purchase committee will be final.

**B. Cancellation of Tender:**

Notwithstanding anything specified in this tender document, IISc Bangalore, in its sole discretion, unconditionally and without having to assign any reason, reserves the rights:

- a. To accept OR reject lowest tender or any other tender or all the tenders.
- b. To accept any tender in full or in part.
- c. To reject the tender, offer not confirming to the tender terms.

**C. Validity of the offer:**

The offer shall be valid 90 Days from the date of opening of the commercial bid.

**D. Evaluation of the offer:**

1. The technical bid (Part A) will be opened first and evaluated.
2. Bidders meeting the required eligibility criteria as stated in Section 2 of this document shall only be considered for Commercial Bid (Part B) opening. Further, agencies not furnishing the documentary evidence as required will not be considered.
3. Pre-qualification of the bidders shall not imply final acceptance of the Commercial Bid. The agency may be rejected at any point during technical evaluation or during commercial evaluation. The decision in regard to acceptance and / or rejection of any offer in part or

full shall be the sole discretion of IISc Bangalore, and decision in this regard shall be binding on the bidders.

4. The award of contract will be subject to acceptance of the terms and conditions stated in this tender.
5. Any offer which deviates from the vital conditions (as illustrated below) of the tender is liable to be rejected:
  - a. Non-submission of complete offers.
  - b. Receipt of bids after due date and time and or by email / fax (unless specified otherwise).
  - c. Receipt of bids in open conditions.
6. In case any BIDDER is silent on any clauses mentioned in these tender documents, IISc Bangalore shall construe that the BIDDER had accepted the clauses as of the tender and no further claim will be entertained.
7. No revision in the terms and conditions quoted in the offer will be entertained after the last date and time fixed for receipt of tenders.
8. Lowest bid will be calculated based on the total price of all items tendered for Basic equipment along with accessories selected for installation, operation, pre-processing and post processing, optional items, recommended spares, warranty, annual maintenance contract.

**E. Pre-requisites:**

The bidder will provide the prerequisite installation requirement of the equipment along with the technical bid.

**F. Warranty:**

The complete system is to be under warranty period of minimum 1 year including free supply of consumables, spare parts, and data analysis software from the date of functional installation. Vendor should include cost of any spares that are expected to be needed during the warranty period, including electronics, subcomponents, and software. If the instrument is found to be defective, it must be replaced or rectified at the cost of the bidder within 30 days from the date of receipt of written communications from IISc Bangalore. If there is any delay in replacement or rectification, the warranty period should be correspondingly extended.

**G. Annual Maintenance Contract (AMC):**

An annual maintenance contract is not required for this purchase.

**H. Purchase Order:**

1. The order will be placed on the bidder whose bid is accepted by IISc Bangalore based on the terms & conditions mentioned in the tender document.
2. The quantity of the items in tender is only indicative. IISc Bangalore reserves the right to increase /decrease the quantity of the items depending on the requirement.
3. If the quality of the product and service provided is not found satisfactory, IISc Bangalore reserves the right to cancel or amend the contract.
4. After the award of the purchase order, the vendor must provide an Order Acknowledgement within 30 days from the receipt of the Purchase Order

**I. Delivery, Installation and Training:**

The bidder shall provide the lead time to delivery, installation and made functional at IISc Bangalore from the date of receipt of purchase order. The system should be delivered, installed, and made functional within 4-6 months from the date of receipt of purchase order. The supply of the items will be considered as effected only on satisfactory installation and inspection of the system and inspection of all the items and features/capabilities tested by the IISc Bangalore. After successful installation and inspection, the date of taking over of entire system by the IISc Bangalore shall be taken as the start of the warranty period. No partial shipment is allowed. The bidder should also arrange for technical training to the local facility technologists and users. The bidder should provide onsite application training for the local facility technologists and users. The bidder should also arrange technical training for the local facility technologists and users.

**J. Payment Terms:**

For INR quotes, 100% payments will be released after completion of delivery, satisfactory installation, and qualification, subject to TDS as per rules. As per GFR no advance payment can be made to domestic vendors. For INR quotes, the price must be on FOR IISc Bangalore basis only.

For quotes in foreign currency, the quote must be on a CIP Bangalore basis. The payment terms for the foreign currency quotes should be mentioned along with the commercial offer.

**K. Statutory Variation:**

Any statutory increase in the taxes and duties subsequent to bidder's offer, if it takes place within the original contractual delivery date, will be borne by IISc, Bangalore subject to the claim being supported by documentary evidence. However, if any decrease takes place the advantage will have to be passed on to IISc Bangalore.

**L. Dispute and Jurisdiction:**

Any legal disputes arising out of any breach of contract pertaining to this tender shall be settled in the court of competent jurisdiction located within the city of Bangalore, India.

**M. General:**

1. All amendments, time extension, clarifications etc., within the period of submission of the tender will be communicated electronically. No extension in the bid due date/time shall be considered on account of delay in receipt of any document(s) by mail.
2. The bidder may furnish any additional information, which is necessary to establish capabilities to successfully complete the envisaged work. It is however, advised not to furnish superfluous information.
3. The bidder may visit the installation site before submission of tender, with prior intimation.
4. Any information furnished by the bidder found to be incorrect, either immediately or at a later date, would render the bidder liable to be debarred from tendering/taking up of work in IISc, Bangalore.



### Technical and commercial specifications

|     | Description              | Specifications                                                                                                                                                                                                                                                                                                                            |
|-----|--------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 1.  | High Speed Homogenizers  | Automated system for cell disruption and particle size reduction                                                                                                                                                                                                                                                                          |
| 2.  | Sample types             | <p>Compatible with:</p> <p>A. Nanoemulsions</p> <p>B. Liposome preparation</p> <p>C. Fine particle de-agglomeration</p> <p>D. Cell disruption of bacteria, mammalian cells and yeast</p> <p>For each of these applications, the bidder should provide data or journal publications clearly showing the feasibility of the instrument.</p> |
| 3.  | Sample volume            | Should be a continuous process with minimum sample volume <15 mL                                                                                                                                                                                                                                                                          |
| 4.  | Re-circulation of sample | There should be a provision to recirculate the output sample back in input for automated cycling of sample                                                                                                                                                                                                                                |
| 5.  | Maximum Pressure         | 30000 PSI or higher                                                                                                                                                                                                                                                                                                                       |
| 6.  | Maximum Flow Rate        | >75 mL/min for water                                                                                                                                                                                                                                                                                                                      |
| 7.  | Operation controls       | All parameters should be digitally controlled                                                                                                                                                                                                                                                                                             |
| 8.  | Pressure compatibility   | The channels and orifice should be less than 200 µm in size and should not change in dimensions with pressure                                                                                                                                                                                                                             |
| 9.  | Sample feed temperature  | Can work in the range of 15 °C to 75 °C                                                                                                                                                                                                                                                                                                   |
| 10. | Solvent compatibility    | Should be compatible with water based solvents. The instrument should also be compatible with upto 40% mixture of following solvents with water: ethanol, Acetone, and Hexane as solvents. Please mention type of material used for any gaskets, o-rings etc to which solvent will come in touch with.                                    |
| 11. | Scale-up                 | All processes should be compatible for scale-up to pilot and/or production volumes for the same pressure range                                                                                                                                                                                                                            |
| 12. | Power Source             | Should be directly usable with standard single phase or three phase power supply. All cords and accessories necessary to start the instrument should be provided with the quote                                                                                                                                                           |
| 13. | Certification            | Equipment should comply with ISO/CE safety certification. Please provide a copy of valid certificate                                                                                                                                                                                                                                      |
| 14. | Customer base            | Similar or identical item must have been supplied to atleast ten central university/institute. Please provide supporting documents for the same and names and contact details of atleast three referees using your instrument at these organizations.                                                                                     |
| 15. | Warranty                 | Minimum 1 year                                                                                                                                                                                                                                                                                                                            |

|     |                           |                                                                                                                                                                                                                                       |
|-----|---------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 16. | Support                   | Must have dedicated service and application support in Bangalore for this instrument. Please provide details of the number of trained personnel in Bangalore who will provide support                                                 |
| 17. | Support                   | The model proposed should not be discontinued by the company. The bidder should provide a certificate that the model is not discontinued and repair parts will be available for atleast another five years from the date of purchase. |
| 18. | Installation and training | The quote should include installation as well as practical training after installation on the instrument for operation and troubleshooting                                                                                            |
| 19. | Breakdown addressing      | Service call within 48 hours of complaint registration                                                                                                                                                                                |

## **Section 5- Technical Bid**

The technical bid should furnish all requirements of the tender along with all annexures in this section and be submitted to

The Chairperson,  
Attn: Prof. Rachit Agarwal  
TCS Smart-X Hub  
Department of Bioengineering,  
Indian Institute of Science,  
Bangalore – 560012, India.

**Annexure-1:****Details of the Bidder**

The bidder must provide the following mandatory information & attach supporting documents wherever mentioned:

**Details of the Bidder**

| Sl. No. | Item                                                                                         | Details |
|---------|----------------------------------------------------------------------------------------------|---------|
| 1       | Name of the bidder                                                                           |         |
| 2       | Nature of Bidder (Attach an attested copy of Certificate of Incorporation/ Partnership Deed) |         |
| 3       | Registration No/ Trade License, (attach attested copy)                                       |         |
| 4       | Registered Office Address                                                                    |         |
| 5       | Address for communication                                                                    |         |
| 6       | Email ID                                                                                     |         |
| 7       | Website                                                                                      |         |
| 8       | Contact person- Name and Designation                                                         |         |
| 9       | Telephone No                                                                                 |         |
| 10      | PAN No. (attach copy)                                                                        |         |
| 11      | GST No. (attach copy)                                                                        |         |

Signature of the bidder

Name

Designation, Seal

Date

**Annexure-2:**

Declaration regarding experience

To,  
The Chairperson,  
Attn: Prof. Rachit Agarwal  
TCS Smart-X Hub  
Department of Bioengineering,  
Indian Institute of Science,  
Bangalore – 560012, India.

Ref: Tender No: XXXXXXXXXX  
Dated: XXXXX

Dear Sir/Madam

I've carefully reviewed the Terms & Conditions in the above-referred tender. I hereby declare that my company/firm has ----- years of experience in supplying and installing the proposed equipment.

(Signature of the Bidder)

Printed Name  
Designation, Seal

Date:

**Annexure-3:**

## Declaration regarding track record

To,  
The Chairperson,  
Attn: Prof. Rachit Agarwal  
TCS Smart-X Hub  
Department of Bioengineering,  
Indian Institute of Science,  
Bangalore – 560012, India.

Ref: Tender No: XXXXXXXX

Dated: XXXXX

Dear Sir/Madam,

I've carefully reviewed the Terms & Conditions in the above-referred tender. I hereby declare that my company/ firm is not currently debarred/blacklisted by any Government / Semi-Government organizations/institutions in India or abroad. I further certify that I'm a competent officer in my company/firm to make this declaration.

Or

I declare the following

| Sl. No. | Country in which the company is Debarred /blacklisted / case is Pending | Blacklisted/debarred by Government / SemiGovernment/Organizations/Institutions | Reason | Since when and for how long |
|---------|-------------------------------------------------------------------------|--------------------------------------------------------------------------------|--------|-----------------------------|
|         |                                                                         |                                                                                |        |                             |

(NOTE: In case the company/firm was blacklisted previously, please provide the details regarding the period for which the company/firm was blacklisted and the reason/s for the same).

Yours faithfully

(Signature of the Bidder)

Name  
Designation, Seal

Date

#### **Annexure – 4:**

Declaration for acceptance of terms and conditions

To,  
The Chairperson,  
Attn: Prof. Rachit Agarwal  
TCS Smart-X Hub  
Department of Bioengineering,  
Indian Institute of Science,  
Bangalore – 560012, India.

Ref: Tender No: XXXXXX  
Dated: XXXX

Dear Sir/Madam,

I've carefully reviewed the Terms & Conditions mentioned in the above-referred tender document. I declare that all the provisions of this tender document are acceptable to my company. I further certify that I'm an authorized signatory of my company and am, therefore, competent to make this declaration.

Yours faithfully,

(Signature of the Bidder)

Name  
Designation, Seal

Date:

## **Annexure – 5:**

Details of items quoted:

- a. Company Name
- b. Product Name
- c. Part / Catalogue number
- d. Product description / main features
- e. Detailed technical specifications
- f. Remarks

Instructions to bidders:

- 1. Bidder should provide technical specifications of the quoted product/s in detail.
- 2. Bidder should attach product brochures along with the technical bid.
- 3. Bidders should clearly indicate compliance or non-compliance with the technical specifications provided in the tender document.



## **Annexure 6 – Support Certification**

Declaration for instrument support

To,  
The Chairperson,  
Attn: Prof. Rachit Agarwal  
TCS Smart-X Hub  
Department of Bioengineering,  
Indian Institute of Science,  
Bangalore – 560012, India.

Ref: Tender No: XXXXXX  
Dated: XXXX

Dear Sir/Madam,

I've carefully reviewed the Terms & Conditions mentioned in the above-referred tender document. I declare that the model quoted has not been discontinued by our organization and the repair parts will be available for atleast another five years. I further certify that I'm an authorized signatory of my company and am, therefore, competent to make this declaration.

Yours faithfully,

(Signature of the Bidder)

Name  
Designation, Seal

Date:

## Section 6 – Commercial Bid

The commercial bid should be furnished with all requirements of the tender with supporting documents as mentioned under:

| S. No. | Description                                                 | Cat. Number | Quantity | Unit Price | Sub Total |
|--------|-------------------------------------------------------------|-------------|----------|------------|-----------|
| 1.     | Essential items noted in the technical specification        |             |          |            |           |
| 1a     | ...(details of essential items)                             |             |          |            |           |
| 1b     | ...                                                         |             |          |            |           |
|        |                                                             |             |          |            |           |
| 2.     | Optional items noted in the technical specification         |             |          |            |           |
| 2a     | ...(details of Optional items)                              |             |          |            |           |
| 2b     | ...                                                         |             |          |            |           |
|        |                                                             |             |          |            |           |
| 3.     | Accessories for operation and installation                  |             |          |            |           |
| 4.     | All Consumables, spares and software to be supplied locally |             |          |            |           |
| 5.     | Warranty (1 year)                                           |             |          |            |           |
| 6.     | Cost of Insurance                                           |             |          |            |           |
| 7.     | Freight on road                                             |             |          |            |           |

Any additional items

| S. No. | Description | Cat. Number | Quantity | Unit Price | Sub Total |
|--------|-------------|-------------|----------|------------|-----------|
|        |             |             |          |            |           |

Addressed to

The Chairperson,  
Attn: Prof. Rachit Agarwal  
TCS Smart-X Hub  
Department of Bioengineering,  
Indian Institute of Science,  
Bangalore – 560012, India.

## **Section 7 – Checklist**

(This should be enclosed with technical bid - Part A)

The following items must be checked before the Bid is submitted:

### **A. Sealed Envelope “A”: Technical Bid**

1. Section 5 - Technical Bid (each page signed by the authorized signatory and sealed) with the below annexures:
  - a. Annexure 1: Bidders details
  - b. Annexure 2: Declaration regarding experience
  - c. Annexure 3: Declaration regarding clean track record
  - d. Annexure 4: Declaration for acceptance of terms and conditions
  - e. Annexure 5: Details of items quoted
  - f. Annexure 6: Support Certification
2. Copy of this tender document duly signed by the authorized signatory on every page and sealed.

### **B. Sealed Envelop “B”: Commercial Bid**

Section 6: Commercial Bid

Your quotation must be submitted in two envelopes: Technical Bid (Envelope A) and Commercial Bid (Envelope B) super scribing on both the envelopes with Tender No. and due date and both of these in sealed covers and put in a bigger cover which should also be sealed and duly super scribed with Tender No., Tender description & Due Date.