



Indian Institute of Science

Department of Bioengineering



BE, TCS Smart-X Hub, Indian Institute of Science, Bengaluru, 560012, India

rachit@iisc.ac.in

<http://be.iisc.ac.in/~rachit>

22-October-2025

Corrigendum

Tender Number: IISC/BE/25-26/05

Tender Title: Supply and Installation of Light Sheet Microscope

Date of publishing Corrigendum: 22-10-2025

Last date of submission: 28-10-2025

This is to inform all prospective bidders that the following point has been corrected under Technical Specifications (page 9), point-11 in the tender document. Kindly read the item description as

11.	Lasers	Minimum three lasers are required. Each laser should be controlled separately. These lasers should be compatible with following dye imaging. DAPI, FITC/GFP and TRITC
-----	--------	--

Rachit Agarwal
Associate Professor
Department of Bioengineering
Indian Institute of Science, Bengaluru – 560012