



CORRIGENDUM

Tender : IISc-Med-2025-26/G-52 dated on 04.02.2026

Date: 17/02/2026

Tender Name : Global tender for [Ultra Low Height Operating Table with Radiolucent Top](#)

1. Corrigendum to tender for Ultra-Low-Height-Operating-Table-with-Radiolucent-Top with reference IISc-Med-2025-26/G-52 dated on 04.02.2026.
2. Except for Points 1, all other terms and conditions of the original tender remain unchanged.

Tender specification		Corrigendum to tender specification	
Sl. No.	Specification	Sl. No.	Specification
D	Table Functions: should at least have the following movements:	D	Table Functions: should at least have the following movements:
1	The operating table should have height adjustment with a minimum height of 550 mm or less from the floor.	1	The operating table should have height adjustment with a minimum height of 595 mm or less from the floor.

IISc Bangalore