

Global Tender for the Supply, Installation and Commissioning of Portable X-Ray Diffraction System

This is a RFQ (Request for Quotation) from the Indian Institute of Science (IISc), Bangalore, for the “Supply, Installation and Commissioning of Portable X-Ray Diffraction System” at the Department of Aerospace Engineering, Indian Institute of Science, Bangalore.

IISc is India’s best research institute and Department of Aerospace Engineering is a research department with the best academic facilities for research related to Aerospace in the world.

1	Section 1	Bid Schedule	
2	Section 2	Eligibility Criteria	As specified by IISc
3	Section 3	Terms and Conditions	As specified by IISc
4	Section 4	Specifications	Technical specifications
5	Section 5	Technical Bid	Annexure 1: Bidder details
			Annexure 2: Declaration regarding experience of bidder
			Annexure 3: Declaration regarding clean track record of the bidder
			Annexure 4: Declaration of acceptance of tender
			Annexure 5: Terms and conditions. Details of item quoted
6	Section 6	Commercial bid	Quotation with Price, Technical specifications of the Equipment

Section 1 - Bid Schedule

1	Tender No	<b>AE/SGK/Corrosion Prognostics-02</b>
2	Tender Date	25/03/2026
3	Item Description	<u>Supply, Installation and Commissioning of Portable X-Ray Diffraction System</u>
4	Tender Type	Two bid system (i) Technical Bid (Part A) (ii) Commercial Bid (Part B)
5	Place of tender submission	The Chairperson Kind Attn: Prof. S Gopalakrishnan Department of Aerospace Engineering (AE), Indian Institute of Science, Bangalore – 560012
6	Last Date & Time for submission of tender	17-04-2026 @ 5:00 PM
7	For further clarification	krishnan@iisc.ac.in

## Section 2 – Eligibility Criteria

### Prequalification criteria:

1. Preference to local bidders is not mandatory.
2. The bidder should sign and submit the declaration for Acceptance of Terms and Conditions as per -Annexure 4.
3. The Bidder must not be blacklisted/banned/suspended or have a record of any service-related dispute with any organization in India or elsewhere. A declaration to this effect has to be given as per Annexure 3.
4. Only the Original Equipment Manufacturer or their authorized representatives across the globe shall participate in the bid
5. The order will be placed only on the bidder who participated in the bid

## Section 3 – Terms and Conditions

### A) Submission of Tender:

1. All documentations in the tender should be in English.
2. Tender should be submitted in two envelopes (two bid system).
  - a. Technical Bid (Part-A) – Technical bid consisting of all technical details and check list for conformance to technical specifications. The proposal should contain a compliance table with 4 columns in addition to the ones in the technical requirements table that has been included with this RFQ below. The compliance table should include all the items in the same order and format. **The first column should describe your compliance in a “Yes” or “No” response. If “No” the second column should state the extent of deviation. The “third” column should state the reasons for the deviation if any.** The fourth column can be used to compare your tool with that of your competitors or provide details as requested in the technical requirements table below (suppliers who include any indication of prices in the technical bid will be automatically disqualified).
  - b. Commercial Bid (Part-B) – Indicating item wise price for the items mentioned in the technical bid, **as per the format of quotation provided in tender**, and other commercial terms and conditions. **The commercial bid should be broken up to the maximum extent possible into separate items with a cost against each (in particular the optional items) to enable better comparison of price for various configurations across the bidders.** As an option, **please provide itemized cost for any *suggested* accessories/add-ons that may enhance the usability, capability, accuracy or reliability of the tool.** Vendors are encouraged to quote for as many add-ons as their tool portfolio permits.
3. The technical bid and price bid should each be placed in separate sealed covers, superscripting on both the envelopes the tender no. and the due date. Both these sealed covers are to be placed in a bigger cover which should also be sealed and duly superscripted with the Tender No, Tender Description & Due Date.
4. The SEALED COVER superscripting tender number / due date should reach Chairperson’s Office, Department of Aerospace Engineering, Indian Institute of Science Bengaluru–560 012, India on or before due date mentioned in the tender notice. In case the due date happens to be a holiday, the tender will be accepted and opened on the next working day. If the quotation cover is not sealed, it will be rejected.
5. All queries are to be addressed to the person identified in “Section 1 – Bid Schedule” of the tender notice.

6. GST/other taxes, levies etc., are to be indicated separately. The BIDDER should mention GST Registration and PAN in the tender document (Indian Bidders only).
7. If price is not quoted in Commercial Bid as per the format provided in tender document the bid is liable to be rejected.
8. The Institute reserves the right to accept or reject any bid and to annul the bidding process and reject all bids at any time prior to the award of contract, without there by incurring any liability to the affected bidder or bidders or any obligation to inform the affected bidder or bidders.
9. The technical proposal must include references of at least 3 previous installations done in India within last 5 years of similar equipment from the equipment manufacturer. Please provide the names and contact addresses of the three independent referees so that the committee can contact them independently to get a reference.
10. Incomplete bids will be summarily rejected.

#### B) Cancellation of Tender:

Notwithstanding anything specified in this tender document, IISc Bangalore, in its sole discretion, unconditionally and without having to assign any reason, reserves the rights:

- a. To accept OR reject lowest tender or any other tender or all the tenders.
- b. To accept any tender in full or in part.
- c. To reject the tender, offer not confirming to the tender terms.

#### C) Validity of the Offer:

The offer shall be valid 90 Days from the date of opening of the commercial bid.

#### D) Evaluation of Offer:

1. The technical bid (Part A) will be opened first and evaluated.
2. Bidders meeting the required eligibility criteria as stated in Section 2 of this document shall only be considered for Commercial Bid (Part B) opening. Further, agencies not furnishing the documentary evidence as required will not be considered.
3. Pre- qualification of the bidders shall not imply final acceptance of the Commercial Bid. The agency may be rejected at any point during technical evaluation or during commercial evaluation. The decision in regard to acceptance and / or rejection of any offer in part or full shall be the sole discretion of IISc Bangalore, and decision in this regard shall be binding on the bidders.

4. The award of the contract will be subject to acceptance of the terms and conditions stated in this tender.
5. Any offer which deviates from the vital conditions (as illustrated below) of the tender is liable to be rejected:
  - a. Non-submission of complete offers.
  - b. Receipt of bids after due date and time and or by email / fax (unless specified otherwise).
  - c. Receipt of bids in open conditions.
6. In case any BIDDER is silent on any clauses mentioned in these tender documents, IISc Bangalore shall construe that the BIDDER had accepted the clauses as of the tender and no further claim will be entertained.
7. No revision of the terms and conditions quoted in the offer will be entertained after the last date and time fixed for receipt of tenders.
8. Lowest bid will be calculated based on the total price of all items tendered for Basic equipment along with accessories selected for installation, operation, preprocessing and post processing, optional items, recommended spares, warranty, annual maintenance contract.

**E) Pre-requisites:**

The bidder will provide the prerequisite installation requirement of the equipment along with the technical bid.

**F) Warranty:**

The complete system must be under a warranty period of 1 year including free support for technical issues in installation and operation of the equipment. Visiting costs, and accessories/spares involved in the repair and replacement of defective components within the warranty period shall be borne by the supplier. If the instrument is defective, it must be replaced or rectified at the bidder's cost within 30 days from the date of receipt of written communications from IISc, Bangalore. If there is any delay in replacement or rectification, the warranty period should be correspondingly extended.

**G) Annual Maintenance Contract:**

An annual / comprehensive maintenance contract for a period of 2 years post warranty may be provided on completion of the warranty period as an option.

**H) Purchase Order:**

1. The order will be placed on the bidder whose bid is accepted by IISc based on the terms & conditions mentioned in the tender document.
2. The quantity of the items in tender is only indicative. IISc, Bangalore reserves the right to increase /decrease the quantity of the items depending on the requirement.

3. If the quality of the product and service provided is not found satisfactory, IISc, Bangalore reserves the right to cancel or amend the contract.

**I) Delivery, Installation and Training:**

The bidder shall provide the lead time to delivery, installation and made functional at IISc, Bangalore from the date of receipt of purchase order. The system should be delivered, installed and made functional within 180 days from the date of receipt of the purchase order. The supply of the items will be considered as effected only on satisfactory installation and inspection of the system and inspection of all the items and features/capabilities tested by the IISc, Bangalore. After successful installation and inspection, the date of taking over of entire system by the IISc, Bangalore shall be taken as the start of the warranty period. No partial shipment is allowed. The bidder should also arrange for technical training for the users.

**J) Payment Terms:**

The payment will be determined after mutual discussions with the successful bidder.

**K) Statutory Variation:**

Any statutory increase in the taxes and duties subsequent to the bidder's offer, if it takes place within the original contractual delivery date, will be borne by IISc, Bangalore subject to the claim being supported by documentary evidence. However, if any decrease takes place the advantage will have to be passed on to IISc, Bangalore.

**L) Disputes and Jurisdiction:**

Any legal disputes arising out of any breach of contract pertaining to this tender shall be settled in the court of competent jurisdiction located within the city of Bangalore, India.

**M) General:**

1. All amendments, time extension, clarifications etc., within the period of submission of the tender will be communicated electronically. No extension of the bid due date/time shall be considered on account of delay in receipt of any document(s) by mail.
2. The bidder may furnish any additional information, which is necessary to establish capabilities to successfully complete the envisaged work. It is, however, advised not to furnish superfluous information.
3. The bidder may visit the installation site before submission of tender, with prior intimation.
4. All imported equipment should be quoted in the currency of the country of origin, and all locally sourced items should be quoted in Indian Rupees.
5. Any information furnished by the bidder found to be incorrect, either immediately or at a later date, would render the bidder liable to be debarred from tendering/taking up of work in IISc, Bangalore.

**Technical specifications**

<b>Sl. No.</b>	<b>Description</b>	<b>Parameter</b>
1	<b>General requirements:</b>	<ul style="list-style-type: none"> <li>a) The system should be suitable for carrying out Residual Stress (RS) measurement using X-ray diffraction technique.</li> <li>b) The equipment is intended to measure Residual Stresses developed in manufactured Components, Castings, Forgings, In-service Components including Bearings, Welded components of similar &amp; dissimilar material combination etc.</li> <li>c) The equipment should be portable and should be capable of measuring Residual Stresses in both Lab and Field conditions.</li> <li>d) The equipment should be fully computerized measuring system to measure the Residual Stress on Aircraft components outer surface areas like Splines, Gear Tooth profile and Radius etc.</li> <li>e) Residual stress measurement shall be by determining change in Inter-planar spacing of Crystal lattice.</li> </ul>
2	<b>The Equipment should comply to the following Standards:</b>	<ul style="list-style-type: none"> <li>a) <b>ASTM E2860</b> – Test Method for Residual Stress Measurement by X-ray Diffraction for Bearing Steels.</li> <li>b) <b>ASTM E915</b> – Test Method for verifying the Alignment of X-ray Diffraction Instrumentation for Residual Stress Measurement.</li> <li>c) <b>ASTM975</b> – X-ray Determination of Retained Austenite in Steel with Near Random Crystallographic Orientation.</li> <li>d) <b>SAE SP-453</b> – Retained Austenite and Its measurement by X-ray Diffraction.</li> <li>e) <b>EN15305-2008</b> – Non Destructive testing –Test Method for Residual stress analysis by X-ray Diffraction.</li> <li>f) <b>ANSI N43.2</b> – Radiation Safety for X-ray Diffraction and Fluorescence Analysis Equipment.</li> <li>g) Safety norms as per <b>AERB (India)</b> guidelines, as applicable.</li> </ul>

3	The test system shall be capable of measuring Residual Stress in components with polycrystalline grain structure and multiphase material made of various crystalline metallic materials and their alloys.	Materials being tested is : a) Aluminium Alloys
4	Following Certified Reference standards shall be supplied along with ISO/IEC 17025 traceability certificate	a) Zero Stress Powder samples for Aluminium, Qty.-1 No. for each material. b) High Stress Standards for Aluminium alloy with known residual stress, Qty.-1 No.
5	<b>Equipment capabilities:</b>	a) The system should be portable and capable of field operation. b) Facility for capturing diffraction profile over seven or more $\psi$ angles within the required angular range for a given $\{hkl\}$ plane. c) Facility for correcting XRD profile data for d) Gain correction to correct for intensity variations caused by detector itself. e) Background corrections for accounting for sloping peak backgrounds. f) Lorentz-polarization-absorption (LPA) for correcting intensity of diffracted beam subjected to additional $\theta$ dependent effects known as Lorentzian and polarization effects. g) Integrated intensity ratios obtained from two detectors should not be more than 3 and facility for corrective measures to be taken for intensity ratios of greater than 1.6 should be available. h) FWHM values obtained from sample should be measured. i) Exposure time should be clearly identifiable for all measurements j) The maximum acceptable error band in stresses measured using equipment should be as follows: <ol style="list-style-type: none"> <li>1. Normal stress error not to exceed 10% of the normal stress value or 35 MPa, whichever is larger</li> <li>2. Normal strain error not to exceed 10% of the normal strain value or 250 ppm, whichever is larger</li> <li>3. Shear stress errors not to exceed 10% of the shear stress value or 35 MPa, whichever is larger</li> <li>4. Shear strain errors not to exceed 10% of the</li> </ol>

		shear strain value or 125 ppm, whichever is larger.
<b>The test system / analyzer should consists of</b>		
6	<b>Control Unit:</b>	<ol style="list-style-type: none"> <li>1. Control unit should house the power supply, Electronics, high voltage generator, Air-cooled mechanism for the tube and all necessary interlocks required for complete safety.</li> <li>2. X-ray tube power supply should be continuously variable within 5-20 kV / 0-2 mA.</li> <li>3. Measurements need to be carried out with 18 kV and 1 mA, or 20 kV and 0.5mA or 10 kV and 2 mA.</li> <li>4. Power rating is universal. 110-220 VAC 50/60 Hz</li> </ol> <p>C. X-ray tube D. Detectors E. Field Stand and F. Control and Analysis Software.</p>
7	<b>Fully computerized Goniometer system with following features:</b>	<ol style="list-style-type: none"> <li>1. Goniometer system shall be mounted on an anchoring system to ensure parallelism.</li> <li>2. Shall have suitable devices for mounting of X-ray tube and detectors and shall be calibrated as per ASTM standards.</li> <li>3. Shall have programmable provision for selection of Psi / Modified chi method (<math>\chi</math>) minimum range of 45° to +45° or more and Psi mode (<math>\psi</math>) minimum range of -45° to +45° or more for measurement in multi exposure <math>\sin^2(\psi)</math> and omega methods.</li> <li>4. Shall have programmable provision to set and operate the system with oscillation angles over a minimum range of 0° to +6°.</li> <li>5. Shall have suitable provision for fixing the distance between the X-ray source and sample surface and to align and set the sample normal to the source.</li> <li>6. The equipment shall have indication for "X-ray on" and shutter open.</li> <li>7. Goniometer shall be provided with manual focus pointer, which enables accurate positioning of the Goniometer in complex geometries. Also, automatic fast focusing system without need of calibration shall be provided. Preferably flexible positioning of the goniometer shall be provided.</li> <li>8. Integrated interlocks for shutter in place and</li> </ol>

		<p>shutter open and X-ray tube thermal protection sensor shall be provided.</p> <p>9. Manual focus pointer for high accuracy 40mm or more focal distances, instrument to sample height positioning.</p> <p>10. Radius <math>R_0 = 44</math> mm or more.</p> <p>11. Automatic fast focusing system without the need of calibration.</p> <p>12. Should accept 16mm X-ray tubes in a cartridge mount for easy X-ray tube changes without re-alignment of Goniometer and to ensure tube changes in less than 5 minutes.</p> <p>13. Control of Goniometer should be through PC/software.</p>
9	<b>C. X-ray Tubes</b>	<ol style="list-style-type: none"> <li>1. X-ray tubes for checking stated materials with suitable K beta Filters with targets of: <ol style="list-style-type: none"> <li>a. Cobalt, Qty-1 Nos.</li> </ol> </li> <li>2. The X-ray tubes shall have a minimum capacity of 20W to provide 10-20KV and 1-2mA current during measurement.</li> <li>3. The tubes shall have Air-cooled Mechanism It should be a low maintenance system.</li> <li>4. The tubes shall be provided with a minimum cable length of 5 meters.</li> <li>5. X-ray beam size apertures – standard round (0.5, 1 and 2mm) and rectangular shape (0.5x3, 1x3, 2x5 mm).</li> <li>6. It should have manual and automated focus pointer.</li> </ol>
10	<b>D. X-ray Detector:</b> Dual (2) photon counting type X-ray detectors with following features:	<ol style="list-style-type: none"> <li>1. 2 detectors should ensure that twice as much data is recorded in a single measurement to enable direct determination of Shear stress.</li> <li>2. Configurable in psi geometry.</li> <li>3. Single exposure, double exposure and multiple exposure <math>\text{Sin}^2 \Psi</math> techniques.</li> <li>4. Detector Width (<math>2\theta</math>) should be <math>20^\circ</math></li> <li>5. Detector peak shift resolution: <math>0.01^\circ 2\theta</math>.</li> <li>6. Fully adjustable Bragg angle position of the detectors.</li> <li>7. Simple and easy detector position adjustment.</li> <li>8. Detectors should be of Zero maintenance and should not degrade with exposure to X-rays.</li> <li>9. No calibration should be required when using in: <ol style="list-style-type: none"> <li>a. "full-arc" <math>170-130^\circ</math> Bragg angle position</li> </ol> </li> </ol>

		<p>and</p> <p>b. "half-arc" 170-150° for when using only one detector.</p>
11	<b>E. Mapping:</b>	<ol style="list-style-type: none"> <li>1. It should have 50-mm linear mapping slide.</li> <li>2. This should include high precision linear slide 50mm travel.</li> <li>3. It should have Goniometer mounting bracket, detector holder bracket, electronics and cabling for control by XRD unit.</li> <li>4. Handheld motion control pendant. 50 mm of travel with 0.01 mm accuracy.</li> </ol>
12	<b>F. Control and Analysis Software:</b>	<ol style="list-style-type: none"> <li>1. The system should be fully computer controlled by Windows (latest version) based Laptop computer. Laptop computer should be part of the system and to be supplied with pre-installed relevant software and office suite.</li> <li>2. Easy to use interface for operation and control of various functions of the analyzer.</li> <li>3. The control should include: setting of X-ray Tube currents, voltage, start up and shut down, mode selection, selection of psi (<math>\psi</math>), modified chi (<math>\chi</math>) and (<math>\phi</math>) phi angles, oscillation magnitude, exposure times, setting of measurement parameters, automatic measurement and post processing.</li> <li>4. The system should be provided with suitable software for data acquisition and data processing with all standard peak shift calculation methods as per latest standards. The data should be displayed on-line graphically during measurement for both the detectors.</li> <li>5. The software should have capability for post processing on off-line mode for different modes of calculation of peak shift methods.</li> <li>6. The software should have provision to compute tri-axial stress, principal stress, stress tensor, shear stress, maximum shear analysis etc. from the measured data.</li> <li>7. The software should have provision for data storage &amp; retrieval in all standard formats including EXCEL &amp; PDF.</li> <li>8. The software should have library data on selection of suitable tubes and <math>2\theta</math> positions for various</li> </ol>

		<p>materials to be analysed.</p> <p>9. The software shall have provision for taking hardcopy of the test results and customization of reports.</p> <p>10. The software shall have provision for Line maps.</p> <p>11. Material removal corrections and depth of penetration correction should be provided.</p> <p>12. Interactive help wizard for faster learning, fully editable constants, libraries and system diagnostics.</p> <p>13. A comprehensive, easy to use, Windows based data collection and stress analysis package which includes Linear and elliptical regression, "d" vs. "sin2Ψ", stress, shear stress, Dolle-Hauk method, Triaxial method. Parabolic, Gaussian, Pearson VII, Cauchy, centroid, centered centroid, mid-chord fitting. Graphical display of "d", "2θ", "intensity", "breadth", "FWHM" vs "sin2ψ" or sin2χ.</p> <p>14. The equipment should be compatible with work of measurement of residual stresses for different materials.</p> <p>15. It should have suitable recording facility to store test data along with date, time, place of testing, sample id and operator's identification.</p> <p>16. Machine and software user interface should be in English only.</p>
13	<b>Safety:</b>	<p>1. The equipment shall meet the requirements of relevant ANSI standards and AERB India guidelines.</p> <p>2. Safety beacon with shutter open and X-ray on warning lights should be provided. It should have long life LED lamps.</p> <p>3. Automatic shut down if shutter gets struck, opened or removed, tube shielding is loose or removed.</p> <p>4. The system shall have indication for "X-ray on".</p> <p>5. Enclosure with door interlocks as per AERB norms.</p> <p>6. Maximum radiation level 1 Micro Sievert/ hour as per AERB norms.</p>
	Spares and Accessories: Along with all standard accessories following to be supplied:	<p>1. Travel Cases: The system should be portable and light weight. Supply should include impact resistant travel cases for the Controller, Goniometer, Field stand and cables suitable for</p>

		<p>transportation by air.</p> <p>2. Standard tool kit should be provided for general maintenance service.</p>
15	<b>Supply, Installation, Commissioning, Prove out and Training:</b>	<ol style="list-style-type: none"> <li>1. Supplier shall be responsible for supply, installation and commissioning of the machine at IISc Bangalore.</li> <li>2. After successful commissioning of the system, prove out trials to be carried out considering actual applications.</li> <li>3. Supplier shall impart training to IISc Bangalore operators and maintenance personnel on operation and maintenance of the machine at IISc Bangalore for minimum 4 working days.</li> <li>4. Training should cover the following aspects: Operation &amp; maintenance of the machine, Safety aspects, PC based system &amp; its operation, Trouble shooting, software application, all special features of the machine, electrical/ mechanical/ electronics system etc.</li> </ol>
16	<b>Final Acceptance:</b>	Final Acceptance of the equipment will be provided after satisfactory completion of supply, Installation, Commissioning, Prove out, Training and Demonstration of the operation of Residual Stress Measurement equipment with the required Aluminium materials by supplier at IISc Bangalore.
18	<b>Documentation:</b> Supplier should provide both hard copy and soft copy of following documents in English along with the equipment:	<ol style="list-style-type: none"> <li>1. Operation and Maintenance manual of complete system and accessories covering operation and maintenance procedures.</li> <li>2. Preventive maintenance check list for Electrical and mechanical system.</li> <li>3. Other relevant documents (If any).</li> </ol>
19	<b>Supplier reference: Customer</b>	Supplier must provide information on supply of similar equipment to other customers.

#### A. Training and demonstration

Training in usage of the machine (hardware and software) must be demonstrated by the successful bidder at bidder's cost to the end users at IISc, Bangalore. The vendor should be responsible for the complete system installation, functioning maintenance, and training by trained personnel. Bidders should clearly specify the after sales service/application support capabilities without any additional cost.

## Section 5 – Technical Bid

The technical bid should furnish all requirements of the tender along with all annexures in this section and submitted to

Chairperson's Office  
Department of Aerospace Engineering  
*Landmark: near D gate*  
Indian Institute of Science Bengaluru – 560 012

### Annexure-1:

#### Details of the Bidder

The bidder must provide the following mandatory information & attach supporting documents wherever mentioned:

#### Details of the Bidder

Sl. No	Items	Details
1.	Name of the Bidder	
2.	Nature of Bidder (Attach attested copy of Certificate of Incorporation/ Partnership Deed)	
3.	Registration No/ Trade License, (attach attested copy)	
4.	Registered Office Address	
5.	Address for communication	
6.	Contact person- Name and Designation	
7.	Telephone No	
8.	Email ID	
9.	Website	
10.	PAN No. (attach copy)	
11.	GST No. (attach copy)	

Signature of the Bidder

Name  
Designation, Seal

Date:

### Annexure-2:

Declaration regarding experience

To,  
Chairperson's Office  
Department of Aerospace Engineering  
*Landmark: near D gate*  
Indian Institute of Science Bengaluru – 560 012

Ref: Tender No: XXXXXXXXXX  
Dated: XXXXX

Supply, Installation, Commissioning and Training of "Portable X-Ray Diffraction System"

Sir,

I've carefully gone through the Terms & Conditions contained in the above referred tender. I hereby declare that my company / firm has ---- years of experience in supplying and installing multimode plate readers.

(Signature of the Bidder)  
Printed Name  
Designation, Seal

Date:

Annexure-3:

Declaration regarding track record

To,  
Chairperson's Office  
Department of Aerospace Engineering  
*Landmark: near D gate*  
Indian Institute of Science Bengaluru – 560 012

Ref: Tender No: XXXXXXXX

Dated: XXXXX

Supply, Installation, Commissioning and Training of “Portable X-Ray Diffraction System”

Sir,

I've carefully gone through the Terms & Conditions contained in the above referred tender. I hereby declare that my company/ firm is not currently debarred / blacklisted by any Government / Semi Government organizations / institutions in India or abroad. I further certify that I'm competent officer in my company / firm to make this declaration.

Or

I declare the following

Sl.No.	Country in which the company is Debarred /blacklisted / case is Pending	Blacklisted / debarred by Government / Semi Government/Organizations /Institutions	Reason	Since when and for how long

(NOTE: In case the company / firm was blacklisted previously, please provide the details regarding the period for which the company / firm was blacklisted and the reason/s for the same).

Yours faithfully  
(Signature of the Bidder)

Name  
Designation, Seal

Date:

Annexure – 4:

Declaration for acceptance of terms and conditions

To,

Chairperson's Office

Department of Aerospace Engineering

*Landmark: near D gate*

Indian Institute of Science Bengaluru – 560 012

Ref: Tender No: XXXXXX

Dated: XXXX

Supply, Installation, Commissioning and Training of “Portable X-Ray Diffraction System”

Sir,

I've carefully gone through the Terms & Conditions as mentioned in the above referred tender document. I declare that all the provisions of this tender document are acceptable to my company. I further certify that I'm an authorized signatory of my company and am, therefore, competent to make this declaration.

Yours faithfully,

(Signature of the Bidder)

Name

Designation, Seal

Date:

Annexure – 5:

Details of items quoted:

- a. Company Name
- b. Product Name
- c. Part / Catalogue number
- d. Product description / main features
- e. Detailed technical specifications
- f. Remarks

Instructions to bidders:

1. Bidder should provide technical specifications of the quoted product/s in detail.
2. Bidder should attach product brochures along with technical bid.
3. Bidders should clearly indicate compliance or non-compliance of the technical specifications provided in the tender document.

Section 6 – Commercial Bid

**The commercial bid should be furnished with all requirements of the tender with supporting documents as mentioned under:**

S.No	Description	Cat. Number	Quantity	Unit Price	Sub total
1.	Essential items noted in the technical specification				
1.a	... (details of essential items)				
1.b	...				
2.	Optional items noted in the technical specification				
2.a	... (details of essential items)				
2.b	...				
3.	Accessories for operation and installation				
4.	All Consumables, spares and software to be supplied locally				
5.	Warranty (1 years)				
6.	AMC / CMC				
7.	Cost of Insurance and Airfreight				
8.	FOR/CIP/CIF IISc, Bengaluru				

Any additional items

S.No	Description	Cat. Number	Quantity	Unit Price	Sub total

Addressed to  
 Chairperson's Office  
 Department of Aerospace Engineering

*Landmark: near D gate*  
Indian Institute of Science Bengaluru – 560 012

## Section 7 – Checklist

**(This should be enclosed with technical bid- Part A)**

**The following items must be checked before the Bid is submitted:**

### 1. Sealed Envelope “A”: Technical Bid

1. **Section 5- Technical Bid (each page signed by the authorized signatory and sealed) with the below annexures:**
  - a. **Annexure 1: Bidders details**
  - b. **Annexure 2: Declaration regarding experience**
  - c. **Annexure 3: Declaration regarding clean track record**
  - d. **Annexure 4: Declaration for acceptance of terms and conditions**
  - e. **Annexure 5: Details of items quoted**
2. **Copy of this tender document duly signed by the authorized signatory on every page and sealed.**

### 2. Sealed Envelope “B”: Commercial Bid

#### **Section 6: Commercial Bid**

**Your quotation must be submitted in two envelopes: Technical Bid (Envelope A) and Commercial Bid (Envelope B) super scribing on both the envelopes with Tender No. and due date and both of these in sealed covers and put in a bigger cover which should also be sealed and duly super scribed with Tender No., Tender description & Due Date.**