

Date: 18.06.26

Tender (Ref: IISc-Med-2026-27/G-22)

GLOBAL TENDER ENQUIRY

To Whom It May Concern

This Request for Quotation (RFQ) invites comprehensive proposals for the design, supply, installation, testing, commissioning, calibration, validation, warranty, and lifecycle support of a next-generation **SPECT-CT system based on advanced Cadmium Zinc Telluride (CZT) detector technology** for the Indian Institute of Science.

The system shall function as a state-of-the-art molecular imaging platform offering high sensitivity, superior energy resolution, rapid acquisition capability, and robust quantitative imaging performance to meet clinical, academic, and translational research requirements.

The scope shall include a fully integrated solution comprising a **CZT-based SPECT gantry**, high-performance digital detector modules with advanced count-processing electronics, a **diagnostic multi-slice CT subsystem** with dose optimization and iterative reconstruction capabilities, dedicated acquisition, processing, and reporting workstations, advanced clinical and research application software, and all essential accessories, quality assurance tools, and radiation safety components required for complete and uninterrupted operation.

The proposed system shall support streamlined digital workflows encompassing patient positioning, protocol-driven acquisition, automated image processing, advanced reconstruction, quantitative analysis, multimodality image fusion, and structured reporting. The system shall be capable of supporting routine and advanced nuclear medicine applications, including but not limited to **cardiology, oncology, neurology, endocrinology, infection and inflammation imaging, theranostics, and emerging quantitative SPECT applications**, while also facilitating academic research and development activities.

The system shall ensure seamless interoperability with existing and future **Hospital Information Systems (HIS), Radiology Information Systems (RIS), Picture Archiving and Communication Systems (PACS), DICOM networks, and institutional IT infrastructure**. The architecture shall be scalable and upgradeable to accommodate future advancements in detector technology, software applications, artificial intelligence-enabled workflows, and quantitative imaging capabilities.

The vendor's scope shall include comprehensive workflow solution for SPECT-CT, Low dose therapy and iodine therapy along with **site planning inputs**, including radiation shielding recommendations, room layout optimization, and utility requirements for SPECT-CT.

The vendor shall be responsible for installation, acceptance testing, performance validation, calibration, statutory and regulatory compliance including applicable **AERB requirements**, preparation and submission of all necessary documentation, and provision of comprehensive application and technical training for users and biomedical engineering personnel in accordance with relevant national and international standards.

Further details about IISc can be referred from:

<https://medicine.IISc.ac.in/>

A. Procedure:

1. Vendors are required to submit a technical proposal and a commercial proposal in two separate sealed envelopes. Only vendors who meet the technical requirement will be considered for commercial negotiation.
2. The deadline for submission of proposals is 9TH **JULY 2026, Thursday, 5:30 pm Indian Standard Time**.
3. Bids in the sealed envelope should arrive at the office of Dean (A & F), Main building, Indian Institute of Science, Bangalore 560012, India, by the above deadline.
4. The technical proposal should contain a technical compliance table with 6 columns.
 - a. The first column must list the technical requirements in the order that they are given in the technical requirement below in tender specifications.
 - b. The second column should provide specifications of the equipment against the requirement (please provide quantitative responses wherever possible.)
 - c. The third column should describe your compliance with a "Yes" or "No" only. Ensure that the entries in column 2 and column 3 are consistent.
 - d. The fourth column should state the reasons/explanations/context for deviations, if any.
 - e. The fifth column can contain additional remarks from the OEM. You can use this opportunity to highlight

- technical features and qualify the response of previous columns.
- f. The Sixth column should contain the datasheet & technical offer Page reference number.
 - g. If the required information is not available in the Product Data Sheet and printed technical literature, it must be authenticated by the competent authority of the principal manufacturer, and in case of any discrepancy, the decision of the Technical Committee shall be final and binding on the supplier; additionally, the vendor must provide a legally binding declaration stating that the required information will be demonstrated at the time of handover and commissioning
5. Vendors are encouraged to highlight the advantages of their equipment over comparable equipment from the competitors.
 6. In the commercial bid, please provide the itemized cost of the equipment and required accessories, etc.
 7. Please provide itemized cost for any suggested/optional accessories/add-on items that may enhance the equipment usability, capability, accuracy or reliability. Vendors are encouraged to quote for as many add-ons as their product portfolio permits.
 8. In the quote, you are requested to provide itemized cost for spares, accessories, and consumables expected over 2 years of use.
 9. Please indicate the warranty provided with the equipment.
 10. Any questions or clarifications can be directed to:

Dean (A & F)
Main building, Indian Institute of Science,
Bangalore 560012
Office@iiscmedicalschoolfoundation.org

B. Terms and Conditions

1. Only the Original Equipment Manufacturer or their authorized representatives across the globe shall participate in the bid.
2. The order will be placed only on the bidder who participated in the bid.
3. The decision of the purchase committee of IISc will be final.
4. The vendor is responsible for the planning, supply, installation, testing and commissioning of the equipment & the training of personnel of the installed equipment at the IISc.
5. The RFQ must include references to previous installations including the list of all customers where similar systems were installed in the past 5 years. Please provide the names and contact addresses of the referees so that the committee can contact them independently. Details of such systems with model numbers and users should be provided. The reference letters can be used to disqualify vendors with poor track records of service, build quality, system performance, or poor availability of spares.
6. The vendor should have qualified technical service personnel for the equipment based in India and must assure a response time of <2 hours after receiving a service request. The schedule for periodic preventive maintenance for the equipment and all the items related to OEMs should be provided.
7. The indenter reserves the right to withhold placement of the final order and to reject all or any of the quotations and to split up the requirements or relax any or all of the above conditions without assigning any reason.
8. Wherever requested in this specifications sheet, data must be supplied along with technical compliance documents. Technical bids without supporting data will be deemed as technically non-compliant.
9. Upon request, all guaranteed specifications will have to be demonstrated in an active installation. Failure to demonstrate any promised specifications will be deemed as technical non-compliance.
10. Printed literature and published papers to support compliance with the prescribed specifications may be provided duly authenticated by qualified personnel in the company.
11. Technical evaluation by the IISc may include a demonstration to verify the functionalities and capabilities of the equipment quoted. Any discrepancy between the promised and demonstrated specifications will be deemed technical non-compliance. If the need arises, the vendor must be ready to visit IISc for a techno commercial discussion physically.
12. The validity of commercial quotations should be at least 90 days from the last date for the submission of tender documents.
13. **Payment terms:** LC will be opened with 70% payment on shipment of the item and remaining 20% on installation, testing & commissioning and 10% on user satisfaction. Insurance coverage should be till the commissioning of equipment.
14. The functionalities and capabilities of the equipment to be provided as part of documentation. Any discrepancy in technical specification between what was committed during technical evaluation and demonstrated specification on ground will be deemed technical non-compliance. If the need arises, the vendor must be ready to visit IISc for a techno commercial discussion in person.

C. Other terms

1. Shipment and Delivery Terms

1.1 Partial Shipments

- a. Partial shipments are allowed; however, transshipment is strictly prohibited.

1.2 Delivery Confirmation

- a. Delivery shall only be made after receiving written confirmation from the IISc purchase team.

1.3 Consignee Details

- a. The address of the consignee and the markings on the containers must be clearly stated as per the details provided by IISc.

1.4 Packing Slip and Documentation

- a. A packing slip detailing each item and its quantity shall accompany every shipment.
- b. The packing slip must be securely attached to the exterior of one of the containers in a visible manner.
- c. The purchase order (PO) number must be clearly marked on all packing slips, invoices, and correspondence.

1.5 Missing Items and Substitutions

- a. Any items that are not found upon delivery must be clearly noted on the packing slip, and the anticipated availability of such items shall be indicated.
- b. Substitutions of items shall not be made without prior written authorization from IISc.

1.6 Packing of Fragile Equipment

- a. Fragile equipment shall be packed in wooden boxes to prevent damage during transit.

1.7 Packing of Critical Components

- a. Critical components must be packed using foam/bubble wrap and cartons, and securely stuffed within containers to prevent any damage during transit or handling at the site.

1.8 Protection during Transit

- a. The Seller shall ensure that all items are securely protected and packed in accordance with best established practices to avoid damage under conditions such as multiple handling, transportation by ship/road, storage, and exposure to heat, moisture, rain, etc.

1.9 Seller's responsibility for damage

- a. The Seller shall bear full responsibility for any breakage, damage, or pilferage (including during transit or handling within the hospital) resulting from faulty packing.

1.10 Marking and Packing Slip

- a. All packages must be visibly marked with the purchase order (PO) number and name of the Buyer in bold letters.
- b. Copies of the packing slip must also be placed inside each package.

2. Insurance and Freight

- a. The cost of all Freight & Insurance is Included in the purchase order value will be arranged by the supplier. The insurance should be from the vendor warehouse to the site till Installation & commissioning at IISc.

2.2 Seller Notification for Insurance

- a. If IISc needs to arrange insurance, the Seller must notify promptly.

3. Warranty Terms

- 3.1** The equipment along with all the 3rd party items should carry a warranty of 12 months from the date of successful commissioning.

3.2 The warranty shall commence from the submission of a duly filled "Medical Equipment Acceptance Sheet Checklist," accompanied by all relevant documents, as per the specifications and requirements.

3.3 After-Sale Service

- a. After-sales service will be provided by a service engineer trained by the principal company.
- b. The credentials and certification of the service engineer shall be shared with IISc for approval.

3.4 Preventive Maintenance and Calibration

- a. Preventive maintenance and calibration shall be performed according to the recommendations of the Original Equipment Manufacturer (OEM).
- b. Preventive maintenance and calibration shall include calibration for any major breakdowns and be conducted in accordance with local rules and regulations, as well as OEM recommendations.
- c. Maintenance and calibration shall also be based on the equipment performance history, using calibrated equipment traceable to international or NABL standards, as required.

3.5 Responsibility for Malfunctions

- a. The seller shall take full responsibility for any mishaps or malfunctions related to the ordered equipment caused by delayed periodic maintenance or calibration under warranty & subsequently in a comprehensive annual maintenance contract.

3.6 Maintenance and Calibration Costs

- a. Preventive maintenance and calibration shall be executed free of cost during the warranty and Annual Maintenance Contract (AMC) period.
- b. The seller shall clearly inform IISc about the list of consumables or maintenance kits that may incur additional costs (not covered under the maintenance contract) before the equipment is supplied.
- c. All accessories, including computer systems, printers, laptops, and software versions, shall be covered free of charge under warranty, rental contracts, and subsequent maintenance agreements.
- d. The vendor shall provide a separate quotation for the one-time maintenance call cost. This cost should cover the technician's visit charge, labor, and basic service expenses for each individual maintenance call requested by the customer (On call charges)

3.7 Annual Maintenance Contract (AMC) and Comprehensive AMC (CAMC)

The AMC & CAMC rate shall be quoted absolute value of the equipment cost per year till nine years post warranty period of equipment. Please refer to the template for GENERAL TERMS AND CONDITIONS FOR COMPREHENSIVE/LABOUR MAINTENANCE CONTRACT(CMC/AMC).

3.8 No Additional Terms to be imposed

- a. The seller shall not impose any additional terms on the buyer when an Annual Maintenance Contract is established on a yearly basis.
- b. All the terms mentioned in the tender and subsequent purchase order shall remain applicable without any modifications.

3.9 Warranty Terms during CAMC

- a. The warranty terms, including those for preventive maintenance and calibration, shall remain valid and applicable throughout the duration of the CAMC, as per the terms outlined in the tender and subsequent purchase order.

3.10 Payment for AMC and CAMC

- a. Payment for CAMC will be made on a quarterly or annual basis.
- b. Payments will be disbursed upon the successful completion of preventive maintenance and calibration activities, in line with the terms and conditions of the tender and subsequent purchase order.

3.11 Consumables List

- a. The vendor shall provide a list of consumables required for the equipment, along with their associated costs, before the supply of the equipment to IISc.

3.12 Equipment Recall and Standby Equipment

- a. The vendor shall notify IISc of any recall related to the supplied equipment and ensure proper action is taken as per the buyer's recall terms and policies.
- b. In the event of an equipment recall, the seller shall provide suitable standby equipment, ensuring the clinical functionality of the buyer is not impacted.
- c. Any open recall or Field Safety Corrective Action (FSCA) associated with the quoted model shall be **fully disclosed** by the bidder in the technical bid submission.

3.13 Adverse Event Reporting

- a. Any adverse events associated with the medical devices shall be promptly reported to IISc.
- b. The vendor shall ensure that any adverse event is communicated to the National Collaboration Centre-Materiovigilance Programme of India, in accordance with regulatory requirements.

4. Maintenance and Calibration

4.1 Preventive Maintenance and Calibration

- a. Preventive maintenance and calibration will be conducted free of charge under the warranty period and any subsequent Annual Maintenance Contracts (AMC).
- b. Calibration will be performed in accordance with industry standards and OEM specifications.

4.2 Report of Maintenance and Calibration

- a. The Seller shall provide a report of maintenance and calibration with details of the work performed, including calibration standards and methods.

4.3 Qualification of Engineers

- a. The Seller must ensure the trained engineers are certified and qualified for preventive maintenance and calibration.

5. Spare Parts

5.1 Supply of Spare Parts

- a. The Seller shall supply spare parts for the entire lifetime of the equipment and guarantee availability for a minimum of 10 years from the date of commissioning of equipment.

5.2 Price of Spare Parts

- a. The Seller will provide the prices of major spare parts, which should not exceed 30% of the total equipment value.
- b. A list of critical spare parts and their estimated prices shall be submitted with the tender as part of commercial bid.

5.3 Spare Parts Availability

- a. The Seller must maintain a minimum stock of spare parts to ensure quick availability for repairs.

5.4 Spare Parts Pricing

- a. The prices of spares shall be firm for 5 years, after which adjustments can be negotiated considering inflation and exchange variations.

6. Uptime and Compensation

6.1 Uptime Requirement

- a. The bidder must ensure a minimum uptime of 98% based on a 365-day working year.
- b. In case the uptime falls below the specified 98%, the Warranty/CAMC shall be extended by a ratio of 1:7 days for every additional day of downtime.

6.2 Compensation for Test Failures or Erroneous Results

- a. The seller shall be liable to compensate the buyer for any test failures or erroneous results generated by the

ordered equipment.

- b. The compensation amount will be mutually agreed upon by both parties, and this provision will be legally binding.

7. Software and Support Services

7.1 Software Licenses

- a. All software supplied as part of the equipment must come with the necessary licenses for use in India.
- b. The seller shall provide a copy of the software license along with proof of ownership.

The supplied application & operating system software will be kept updated in the form of Free of cost as & when they are released by the factory.

However, for new application software any additional hardware is needed, the cost will be borne by IISc management at negotiated special price.

7.2 Software Support Services

- a. Any software updates or bug-fixing services will be free of charge during the lifetime of equipment.

8. Integration with Clients HIS & PACS-RIS

8.1 Integration Requirement

- a. The Seller must integrate the equipment with clients' Hospital Information System (HIS) & PACS-RIS at no extra cost, as applicable.

9. Confidentiality and Ownership Transfer

9.1 Confidentiality

- a. The service provider must not acquire or retain any confidential data from IISc.

9.2 Ownership Transfer

- a. Any change in the ownership of the principal company must honor all existing agreements with IISc.

10. Recall of Equipment

10.1 Equipment Recall

- a. In the event of any recall of equipment, the Seller shall promptly inform IISc in writing.
- b. During the period when the equipment is under recall, the Seller shall provide suitable standby equipment of similar or higher specifications to IISc, at no cost.

11. Force Majeure

If either Party is unable to carry out his obligations under this Contract due to an Act of God, war, riot, blockade, strike (i.e. national/ state or city), lockout, flood or earthquake or Government orders/ restrictions not within the control of the parties hereto which results in an Inability, in spite of due diligence of either party in performing its obligation. In time, this Contract shall remain effective, but the obligation which the affected party is unable to carry out shall be suspended for a period equal to the duration of the relevant circumstances provided that :

The non-performing party shall give the other Party prior written notice describing particulars of the Inability including but not limited to the nature of occurrence with its expected duration and the steps which the non-performing parties is taking to fulfil its obligation.

Upon receipt of such notice the other party shall discuss the matter with the non-performing party with a view to helping the non-performing party to fulfil obligations. This clause does not envisage financial assistance.

If in any event the Force Majeure situation continues for a period of three weeks both the parties shall meet again and discuss whether the Contract can be amended to overcome the Force Majeure situation so the Project can proceed further.

Notwithstanding anything contained to the contrary it is clarified that economic hardship, non-availability of material, labour and transport shall not constitute Force Majeure. The overall responsibilities and obligations of the parties shall not be excused by reasons of Force Majeure situation.

Notwithstanding the above If the Force Majeure continues for a period of three months or more in that event without prejudice to the rights of the parties, the Buyer shall have the right thereafter to terminate this contract.

12. Seller's Personnel at Buyer's Premises

12.1 Adherence to Safety Regulations

- a. Seller's personnel on IISc premises must adhere to all IISc safety regulations and protocols.

12.2 Seller's Responsibility for Personnel's Safety

- a. The Seller is responsible for their personnel's safety and health while on IISc premises and shall indemnify IISc for any accidents or injuries.

13. Site Evaluation

- a. The Seller must conduct a site evaluation including transportation path, power, air conditioning and other requirements before equipment installation, as applicable.
- b. The Seller shall submit detailed drawings, specifications, and colour codes for all ordered items for Buyer review and approval via email or other methods, as applicable. Manufacturing shall commence only after drawing approval and joint inspection of the proposed site.

14. Skilled & trained Engineer for Installation

- a. Installation must be carried out by a skilled engineer and is considered complete only when the equipment is fully operational as per the tender specification.

15. Inspection and Quality Plan

15.1 New Equipment Requirement

- a. Only brand-new equipment will be accepted, and it must be accompanied by quality conformance and manufacturer test certificates.

15.2 Training

- a. Hands-on training for IISc engineers and technicians must be provided at no extra cost.

16. Marketing Support

- a. The Seller shall provide marketing support to IISc as mutually agreed upon.

17. Other terms and conditions

- a. **Software Compatibility** – If the equipment includes software, it must support integration with hospital EMR/HIS via HL7/FHIR standards, and required middleware as applicable.
- b. **Standard Accessories** – The system should come with all essential accessories (e.g., power cords, consumables) required for full functionality.
- c. **Regulatory Compliance** – The software should comply with National Health Stack requirements and undergo ABDM sandbox testing, if applicable.
- d. **Data Protection** – Any patient-related data generated by the equipment must adhere to DPDP Act guidelines, if applicable.
- e. **BMS Integration** – The system should include a portal for Building Management System (BMS) integration at no additional cost, if applicable.
- f. **Local Service Support** – Supplier must have a registered office, trained engineers, spare parts, calibration equipment, and installation references in Bangalore.
- g. **Country of Origin Restrictions** – Equipment/materials originating from countries sharing a land border with India will not be accepted.
- h. **Cloud Based facilities**- All cloud-based facilities should be hosted in the IISc by the vendor
- i. Vendor shall provide regulatory certificates (like **CDSO/CE/FDA/ISO/AERB/BIS**) type approval where

applicable) for the quoted model and the same is to be enclosed on the technical bid.

18. Vendor is to provide compliance with remarks against all terms and conditions

- a. The First column should describe your compliance with a “Yes” or “No” only. Ensure that the entries in column 1 and column 2 are consistent.
- b. The Second column should state the reasons/explanations/context for deviations, if any.
- c. The Third column can contain additional remarks from the OEM

19. A soft copy of the technical compliance sheet (only) in both pdf and worksheet like excel format should be submitted in pen drive along with technical bid

TEMPLATE FOR ACCEPTANCE OF MEDICAL EQUIPMENT FOR CLINICAL USAGE

SN	MEDICAL EQUIPMENT PRE-COMMISSIONING CHECK-LIST	Vendor to fill the details
1	Equipment name	
2	Main Unit Model & Serial No	
3	Date of receipt of equipment at site	
4	Goods opening report (item wise)	
5	Principal Company name	
6	Dealer/ Vendor name	
7	Vendor contact details including email address	
8	Equipment Model name	
9	User Department name	
10	End User (Head of Dept) Signature	
11	Clinical Engineers name	
12	Clinical Engineers Signature	
13	Service Engineers name and Contact number	
14	Application specialist name and contact number	
15	Main Unit - hardware as per Purchase Order (Vendor-signed PO and list of items supplied as per PO with invoice) to be enclosed as part of the commissioning documentation.	
16	Main Unit - software as per Purchase Order (Vendor-signed PO and list of software supplied as per PO with invoice) to be enclosed as part of the commissioning documentation.	
17	OEM items as per Purchase Order (Vendor-signed PO and list of items supplied as per PO with invoice) to be enclosed as part of the commissioning documentation.	
18	Accessories as per Purchase Order (Vendor-signed PO and list of items supplied as per PO with invoice) to be enclosed as part of the commissioning documentation.	
19	Consumables as per Purchase order- (Vendor signed PO and List of items supplied as per PO with invoiced) to be enclosed as part of commissioning documentation	
20	Brochure of equipment to be enclosed as part of the commissioning documentation.	
21	Technical Data Sheet to be enclosed as part of the commissioning documentation.	
22	Set of service manuals (1 hard copy & 1 PDF soft copy) to be handed over to the Clinical Engineering Dept.	
23	Set of instruction manuals - Two copies (1 hard copy and 1 PDF) to be handed over to the Clinical Engineering Dept.	

24	List of spares & additional accessories with re-ordering codes and costs used along with the equipment as a standard package (PDF).	
25	Equipment demo training information materials like PPT/Video to be handed over to the Clinical Engineering department.	
26	Duly signed letter from the vendor organization head (MD/CEO) stating that the supplied unit, accessories & OEM items are brand new from the factory, to be enclosed as part of the commissioning documentation.	
27	Quality test certificate of equipment from the factory, duly signed by the factory production in-charge, to be enclosed as part of the commissioning documentation.	
28	Software license document (PDF); including OS, system and application software, and commitment to support over the lifetime of the equipment, to be enclosed as part of the commissioning documentation.	
29	All cables from the equipment should have proper cable management, i.e., cable labeling.	
30	2S and HIRA (Hazard Identification and Risk Assessment) to be conducted during preventive maintenance wherever applicable to keep the working area clean.	
31	First-level training to Clinical Engineering (training certificate).	
32	Application training to the end-user on all functions demonstrated (training certificate).	
33	Do's and Don'ts for the equipment for the user group to be provided as part of the training module, to be enclosed as part of the commissioning documentation.	
34	Preventive maintenance frequency calculated based on Equipment Risk Classification, Usage and Operational Intensity, Manufacturer's Recommendations, Historical Performance, and Failure Data.	
35	Preventive maintenance (PM) checklist to be predefined & duly filled during preventive maintenance, to be enclosed as part of the commissioning documentation.	
36	Preventive maintenance kit specification & details to be shared in advance, to be enclosed as part of the commissioning documentation.	
37	Preventive maintenance schedule should be done during non-clinical work operational hours based on prior approval from the user.	
38	Calibration schedules should be based on Manufacturer's Recommendations and after every major equipment breakdown servicing.	
39	The calibration process should follow NABL 126 guidelines.	
40	With each maintenance work, the service provider should hand over two physical copies of the service report (one for the user and one for the Clinical Engineering Dept.) along with a duly filled PM checklist. If physical copies are not available, soft copies should be provided to both the user and the Clinical Engineering Dept. Accepted downtime in hours & accepted equipment breakdown frequency as per PO terms should be understood by the service team, including downtime penalty	
41	Accepted Downtime in hours & accepted equipment breakdown frequency as per PO terms are understood by the service team including downtime time penalty.	
42	The service provider should maintain a logbook of maintenance at the user site.	
43	Shelf-life details of critical spares/accessories/consumables to be provided, to be enclosed as part of the commissioning documentation.	
44	Commissioning report should include (IQ/PQ/OQ) as part of equipment commissioning documents, duly signed by the user group, to be enclosed as part of the commissioning documentation.	
45	Cleaning & disinfection methodology, including the material used, to be provided at the time of commissioning of equipment, to be enclosed as part of the commissioning documentation.	
46	User application training schedule to be provided along with the PM schedule.	
47	Training materials soft copy (PPT/Video) to be shared for installation sign-off.	
48	Letter from the principal manufacturer stating their commitment to IISc for support of equipment for the coming years as per Purchase Order terms to be provided.	
49	CE/FDA, CDSCO Certificate to be enclosed as part of the commissioning documentation.	
50	The single-phase power cord supplied along with the equipment should have a 3-pin plug (Neutral, Phase, Earth) for Indian usage.	
51	Warranty card and details of the warranty to be enclosed as part of the commissioning documentation.	

52	Short shipped items (if any) with quantity. The warranty will start only after full supply, installation, testing, and commissioning of hardware, application software, and third-party equipment supplied along with the main equipment.	
53	OEM and Dealer Sales and Service Escalation contact details, including CEO/MD, to be enclosed as part of the commissioning documentation.	
54	Life of the equipment as committed during technical discussions to be provided with maintenance and spare support during the course of the year, irrespective of dealer change, as per PO terms and conditions, to be given on the OEM letterhead. In case the OEM stops service support during the sales-committed life, the vendor is expected to compensate with the depreciated cost of equipment or provide buyback or upgrade options according to the hospital's requirements.	
55	Any adverse events and recalls related to the equipment, if reported, need to be intimated to IMSF in a timely manner to ensure patient & staff safety by the vendor.	
	Signature: User Dept Head Head-Clinical Engineering	
	Date and Time	
	All these details should be given in a spiral bound document by vendor to IISc.	
	EQUIPMENT WARRANTY WILL START ONLY AFTER FULL COMPLIANCE OF ABOVE FORM	

GENERAL TERMS AND CONDITIONS FOR COMPREHENSIVE/LABOUR MAINTENANCE CONTRACT(CMC/AMC)	
1)	ALL TERMS AND CONDITIONS REMAIN UNCHANGED AS PER SALES PO
2)	AMC & CMC VALID FROM _____ TO _____
3)	THIS CONTRACT INCLUDES
1	All equipment and items supplied by the OEM are covered under service contracts and must be replaced free of cost under CMC.
2	All equipment must be serviced by trained, authorized service engineers. The training certificate of the engineer must be submitted to the IMSF Clinical Engineering Team in advance.
3	Preventive maintenance frequency is calculated based on equipment risk classification, usage, operational intensity, manufacturer's recommendations, historical performance, and failure data.
4	The equipment preventive maintenance must be performed according to the predefined checklist provided in the service manual.
5	Operating system and anti-virus updates are an integral part of preventive maintenance.
6	The vendor will not allow their service engineer to train junior staff on our equipment.
7	Vendor to attend unlimited breakdown calls.
8	Call response time of two hours to be maintained; response time to attend calls within 2 hours is applicable, including holidays and non-working hours.
9	Breakdown frequency should not exceed twice the frequency of preventive maintenance.
10	Vendor must submit soft copies of all reports in two copies.
11	Vendor must maintain a service logbook at the user department.
12	Yearly downtime and breakdown frequency will be calculated based on the call logbook.
13	Any damage to hospital property during maintenance by the company engineer should be compensated to the hospital.
14	Vendor must ensure two preventive maintenance visits per year before the due date. Any malfunction or harm to the patient due to delayed preventive maintenance or calibration will be the sole responsibility of the vendor, including legal compensation. Preventive maintenance and calibration must be mandatory after repair or replacement of any

	spare parts, and necessary kits are to be provided FOC.
15	A copy of the preventive maintenance report with a checklist and a soft copy of calibration, if applicable, is to be shared within one day of execution. The preventive maintenance and calibration label, with done and due dates, must be affixed to the machine without fail, along with the clinical engineer.
16	Periodic training to clinical engineers and end-users, as and when applicable, is mandatory. Training documents must be provided for all concerned staff prior to the renewal of the contract. It is the vendor's responsibility to ensure training, including application training for all staff, without fail. Training materials (PPT/Video) must be submitted to the clinical engineering team prior to any training.
17	Vendor should provide the cleaning and disinfection protocol for the equipment, carry out necessary training periodically, and ensure that all concerned members are trained on the same.
18	Any recall related to the above equipment must be notified in writing, and required corrective actions must be carried out FOC. Necessary training must be provided to concerned staff.
19	Any adverse event reported must be intimated to the Materiovigilance department, and corrective action must be shared within one working day with the hospital.
20	Complete breakdown details, including downtime and preventive maintenance/calibration history, must be shared before the renewal of the next contract. Any downtime of more than 48 hours must include root cause analysis and corrective & preventive action with due diligence. Service reports must be legible and include call received, call attended, and call closed (including date & time) accurately. Any report missing this information will be deemed incomplete.
21	Unlimited spare support must be provided, except for consumables (filters). All accessories and parts are covered and included in the contract. Spares must be ordered and moved immediately after diagnosis, including during holidays and non-working hours.
22	Uptime must be maintained at 98%, including holidays and non-working hours.
23	Uptime is defined by the machine working for its intended purpose without compromising patient care or revenue. Any deviation will count as downtime, and for any additional downtime, the contract will be extended by 1:7 days.
24	A maximum of two breakdowns per preventive maintenance frequency is permitted. Any deviation will increase the preventive maintenance frequency in the subsequent year with any cost escalation.
25	Standby equipment must be provided within a day if the issue cannot be resolved for movable equipment.
26	The vendor escalation matrix, including sales and service contact details (mobile numbers & email IDs), must be provided without fail.
27	First-level service training must be provided for the concerned equipment, and the training certificate must be provided to the clinical engineering team members.
28	Preventive maintenance must not be executed during peak working hours and must be carried out as per the user's convenience. The preventive maintenance kit is included in the CMC and must be replaced during preventive maintenance.
29	The AMC bill will only be cleared after the submission of the equipment log report, which must include details of downtime and preventive maintenance (PM) or calibration history. This report must be provided prior to the renewal of the contract.
30	For equipment under AMC, the quotation for spare parts must be provided within one day of the service engineer's recommendation in the service report.
31	For equipment under AMC, no cannibalization of spare parts from working equipment by the service engineer is allowed.
32	Any spare part ordered for equipment under CMC must reach the hospital site within 72 hours.
33	All defective spare parts under AMC will be retained by the hospital. For equipment under CAMC, IMSF will mark the spare part as defective, and a non-returnable gate pass will be issued.

Template for purchase order terms

General: Acceptance of this Purchase/ Work Order (hereinafter referred to as "PO/Order") includes the acceptance of the following terms & conditions and is made expressly conditional on Seller's assent to the exact terms contained herein. None

of the terms in the Order may be modified, added to, or superseded, except with the written consent of Indian Institute of Science ("Buyer").

1.Price: The prices mentioned in this Order are the prices at which Buyer has agreed to purchase the Goods or Services (as applicable). No escalation in the aforesaid prices shall be binding on Buyer, notwithstanding anything that may be mentioned in Seller's terms of acceptance of Order.

2.Advice of Dispatch: A full and comprehensive dispatch advice notice shall be sent to stores or concerned departments of the Buyer ("Buyer Stores"). Instructions regarding dispatch & Insurance as mentioned in this Order should be complied with and the packing slips giving reference of Buyer order number shall be included securely with the goods in closed envelopes.

3.Delivery Terms:

(a) Deliver Date: Time is the essence in any Purchase Contract. Time of delivery/performance as mentioned in this Order shall be the essence of the Agreement and no variations shall be permitted except with prior authorization in writing from the Buyer.

(b) Place of Delivery: The goods/services shall be delivered/performed strictly as per the instructions in the Order. All Goods/Services delivered/performed should reach Buyer Stores before 2.00 p.m. on weekdays except that no deliveries/ dispatches shall be made or accepted on Sundays or holidays in the working place of the Buyer.

(c) Delayed Delivery: The time and date of delivery/performance as stipulated in the Order shall be deemed to be the essence of the Agreement. In case of delay in performance of its obligations by the Seller, or any extension granted by the Buyer, the Buyer shall at his option either (i) accept delayed deliveries at price reduced by a sum/ percentage (%) mentioned in the Purchase Order for every week of delay or part thereof; and/or (ii) cancel the Order in part or in full and purchase such cancelled quantities from open market at the prevailing market price at the risk & cost of the Seller without prejudice to his rights under 3(c) (i) noted above in respect to the goods delivered; and/or (iii) refuse to accept the Goods delivered beyond the delivery date and claim/set-off the difference between the prevailing market price and contracted price of such quantity delivered belatedly by the Seller.

(d) Delay due to force majeure: In the event of cause of force majeure occurring within the agreed delivery terms, the delivery date may be extended by the Buyer at its sole and absolute discretion on receipt of application from the Seller without imposition of liquidated damages. Only those cause which have duration of more than seven (7) consecutive calendar days will be considered the cause of force majeure. The Seller must inform the Buyer, by a Registered Post or courier letter duly Certified by the Chamber of Commerce or Statutory Authorities, the beginning and the end of the cause of delay immediately, but in no case later than ten (10) days from the beginning and end of each cause of force majeure as defined above.

(e) The goods shall correspond with the description of the samples of the original specification thereof in full details and must be delivered and dispatched within the stipulated time, as the case may be. Otherwise, the same shall be liable to be rejected and the Seller shall be deemed to have failed to deliver the goods in breach of the PO. The Buyer shall in that event at its sole and absolute discretion, will be entitled to either purchase such goods from other sources on Seller's account, in which case, the Seller shall be liable to pay to the Buyer any difference between the price at which such goods have been purchased and the price calculated at the rate set out in this Order or to hold the Seller liable to pay the Buyer damages for non-delivery of goods for such breach.

(f) Packing: Goods supplied against this order must be suitably and properly packed (conforming to special conditions stipulated by the Buyer, if any, for safe and/or undamaged transport by road or rail.)

4. Examination of goods: Irrespective of the fact that the goods are delivered to the Buyer by the Seller at the Seller's place or at Buyer's said office or are dispatched as per Buyer's instructions by rail or road, the goods shall always be supplied, subject to detailed inspection, at the Buyer works or such other destinations as specified in the Order for ascertaining whether the goods are in conformity with the Agreement or not and until then in no event the Buyer shall be deemed to have accepted such goods and upon any rejection of goods in question the Seller shall be deemed to have failed to deliver the concerned goods in accordance with the Agreement.

5. Rejection/ Removal of rejected goods and replacement: Buyer shall have the right to reject the goods whether in full or parts which are not delivered in accordance with the terms of the PO. within fifteen days from the receipt of the intimation from the Buyer of his rejection to accept the goods the Seller shall remove, at his own cost, the rejected goods from the Buyer's works or wherever such goods are lying. The Buyer shall not be in any way responsible for or be held liable for any loss or deterioration of the rejected goods as this shall be at the Seller's risk entirely. The Seller shall pay to the Buyer reasonable storage charges for storing such rejected goods for a period exceeding 15 days as aforesaid. Upon rejection, if the Seller fails to replace the goods with the goods acceptable to the Buyer within the contractual period then the Buyer may, solely at his discretion, exercise all or any of the following options in respect of the rejected/undelivered quantity:- a. Dispose-off the rejected goods and claim/set-off the difference between the prevailing market price and contracted price of such undelivered/rejected quantity to the Seller's account; and/or b. purchase such undelivered/rejected quantity from the open market at the prevailing market price at the risk and cost of the Seller.

6. Transit Insurance: In case insurance is not included in Seller's scope he must furnish details such as reference, Lorry Receipt, Note No., nature of packing, number of cases, gross weight net weight, train carrying the goods, value of the goods dispatched etc. immediately on dispatch to Buyer's office to take up insurance in case of goods sent by Regd... Post, the Regd. Post parcel No. should be furnished to the Buyer with a packing slip when action will be taken to insure the goods. This procedure will be adopted unless specially advised by the Buyer to the contrary.

7. Insurance: Seller agrees that during the term of its performance hereunder, it shall, at its sole cost, maintain worker's compensation insurance and other legally required insurance in accordance with and meeting requirements of applicable law.

8. Invoices: All bills/ invoices for supplies/ services made bearing registration number of the Seller should be marked to concerned Office or as mentioned in Order (quadruplicate) duly endorsed with Purchase Order, Reference Number and Date

and be accompanied by advice of dispatch detailed packing list and by an appropriate certificate necessary under the GST Registration Rules and Regulations.

9. Billing Instructions: Seller must follow the billing instructions carefully and correctly to enable early settlement of his dues. Disregarding the same may involve delay in such settlement. Seller must mention the following information in his bill: (1) Vendor Code Number (2) Purchase Order Item Number (3) Material Code Number, if any. The abovementioned information will be always available in this Order sent to him. One copy of the above document is to be sent to Buyer at The Assistant Registrar, Stores and Purchase Section, Indian Institute of Science, Sir C V Raman Avenue, Bengaluru-560012 or to the address as advised by the Buyer.

10. Compliance with laws: It is clearly reiterated that the Seller is representing an Entity which is strictly complying with all the Laws of the Land as is expected generally from a Seller of a product. It is also made explicitly clear that (a) the Seller has and shall maintain as valid shall under this order strictly comply with the specifications and the requirements agreed upon. At any given point of time, the seller is obliged to produce all applicable licenses, permits, approvals, authorizations and/or or other statutory approvals required to perform its obligation/s under the PO; (b) shall at all times duly observe, perform and comply with all obligations, requirements and/ or prohibitions contained in any statutes, regulations or ordinance of any authority whether governmental or provincial, relating to or in any way affecting or regulating the respective performance of the PO by it.

11. Standard GST Clause: a. The price quoted in this PO for supply of goods shall be exclusive of any applicable Goods and Services Tax, Customs duties, or any other indirect tax as may be imposed by the Government of India from time to time. The Seller shall provide a proper invoice in the form and manner prescribed under GST Invoice Rules containing all the particulars mentioned therein. In the event that the Seller fails to provide the invoice in the form and manner prescribed under rules, Buyer shall not be liable to make any payment against such invoice. Notwithstanding anything contained anywhere in the Agreement, in the event that the input tax credit of the GST charged by Seller is denied by the tax authorities to Buyer, Buyer shall be entitled to recover such amount from the Seller by way of adjustment from the next invoice. In addition to the amount of GST, Buyer shall also be entitled to recover interest at the applicable rate and penalty, in case any penalty is imposed by the tax authorities on Buyer. b. As required by any applicable legislation, where identifiable cost savings are realised by virtue of the enactment of the GST law, those cost savings will be reflected in the calculations of the consideration under this Agreement and shall be passed on by the Seller to Buyer. c. Event of default clause – In the event that the Seller does not deposit the GST charged on the invoice issued to Buyer or such GST charged on the invoice and paid by Buyer is not reflected in online tax credit ledger on common GSTN portal of the govt. as eligible input tax credit for any reason whatsoever, this Agreement shall be liable to be terminated with immediate effect and Seller shall be liable to pay such damages as may be reasonably estimated by Buyer. In the event that the compliance rating prescribed under the GST Act, 2017 read with GST Rules, 2017 of Seller falls below prescribed level for any reason whatsoever, this Agreement shall be liable to be terminated with immediate effect and Seller shall be liable to pay such damages as may be reasonably estimated by Buyer. d. Representation and warranties clause – The Seller represents and warrants that it shall have and maintain in effect level of compliance rating as prescribed by the govt.

12. Warranty: The Seller warrants that goods and/or services supplier shall be of the highest grade and quality unless otherwise specified; shall conform to the specifications, drawings, samples or other descriptions contained in the Order or furnished or specified by the Buyer; shall be performed in a workmanlike manner; shall be fit and sufficient for the purpose intended; shall not violate any third party intellectual property rights and shall be merchantable, of good material and workmanship and free from all the defects whether latent or patent. In case the same is found to be defective, inter-alia, in respect of materials, workmanship, design or process of manufacturing within a period 12 months after the same had been put in use or 20 months from the date of acceptance of the goods by the Buyer, whichever is earlier, the Seller shall refund the price paid by the Buyer in respect of the said goods. The Seller shall guarantee that the material Seller further agrees that all materials / goods shall be repaired or replaced as the case may be as noted in Clause 20 below. All spare parts should carry the following: a) Name of the Machine b) OEM/Party's name c) Sr. at his / her own expense. The Seller shall be liable for all costs and damages and replacements at the sole option of the Buyer. These warranties are in addition to those implied by or available at law to Purchaser and shall exist notwithstanding the acceptance and/or inspection by Purchaser of all or part of the goods or services.

13. Right of the Buyer to Set Off: In the event, the Seller fails to deliver the goods in accordance with the terms of this PO, the Buyer shall have the right to cancel the PO forthwith and claim refund of any payment made by the Buyer as advance or otherwise to the Seller under the PO. The Buyer shall also have the absolute right to withhold, adjust, and/ or set-off any payment required to be made by the Buyer to the Seller under this PO or any other PO entered into between the parties against the cost, losses, damages etc. suffered by the Buyer due to the failure of the Seller to deliver the Goods in accordance with the terms of this PO, and the Seller expressly waives any objections it may have in this respect.

14. Cancellation/Termination: The Buyer reserves the right to cancel/terminate this Purchase Order or any part thereof. The Buyer shall be entitled to rescind the Agreement wholly or in part in a written notice to the Seller if (i) The Seller fails to comply with the terms of the Purchase Order; or (ii) The Seller goes bankrupt or goes into liquidation proceedings; or (iii) The Seller fails to deliver the goods on time and / or replace the rejected goods promptly; or (iv) the Seller fails to deliver the Goods/Services of desired quality, weight, specification, drawing, layout, design, etc.; or (v) The Seller makes general assignment for the benefit of the creditors; or (vi) Receiver is appointed in respect of property of the Seller. The Buyer shall also be entitled to cancel this Order without assigning any reasons or becoming any way liable in such cancellation.

15. No Assignment: This Purchase Order shall not be assigned to any other agency by the Seller without obtaining prior written consent of the Buyer.

16. Force Majeure: Failure or omission to carry out or observe any of the stipulation or condition of the Agreement shall not give rise to any claim or be deemed a breach of the Agreement if the same shall arise from any of the following causes. viz. the imposition or restriction on Import, Acts of God. The Seller submits his acceptance of this agreement with the above

<p>conditions by acceptance of Buyer's Order even in cases where the confirmation has been made under assumption of different conditions.</p>
<p>17. Special Conditions: Seller will ensure that all statutes, regulations of the Central or State Government are strictly followed. Buyer shall not be liable to pay any damages/compensation due to non-compliance of these rules / regulations by Seller.</p>
<p>18. Arbitration: Any dispute arising out of or in connection with the agreement shall be settled by Arbitration in accordance with the Arbitration Conciliation Act, 1996. The arbitration proceedings shall be conducted in English in Bengaluru by the sole arbitrator appointed by the Buyer. The cost of arbitration shall be shared equally between the parties unless decided otherwise by the arbitrator.</p>
<p>19. Dispute & Jurisdiction of Bengaluru: All disputes shall be subjected to the exclusive jurisdiction of the court in Bengaluru only or as provided in the PO/Order.</p>
<p>20. Limitation of Liability: In no event shall Buyer be liable to Seller, or to Seller's officers, employees or representatives, or to any third party, for any indirect, consequential, incidental, special, punitive or exemplary damages of whatsoever nature (including, but not limited to, lost business, lost profits, damage to goodwill or reputation and/or degradation in value of brands, trademarks or trade names, service names or service marks, or injury to persons) whether arising out of breach of contract, warranty, tort (including negligence, failure to warn or strict liability), contribution, indemnity, subrogation or otherwise.</p>
<p>21. All spare parts should carry the following: a) Name of the equipment b) OEM/Party's name c) Sr. No. as per the catalogue d) Buyer's Order No. and date and e) Quantity all relevant information.</p>
<p>22. Works carried out in Buyer's Institution or premises by the Sellers representatives etc.: Agent representative or employees of the Seller who in pursuance of the Agreement have to work in Buyer/Owner's Institution/Premises will be subject to the rules and regulations existing in the works. The Buyer shall not be liable for any accident which may cause to the Sellers personnel.</p>
<p>23. Intellectual Property Rights: All drawings, specifications and other documents furnished by Buyer and the Buyer's consultants, and copies thereof furnished to the Seller, are for use solely with respect to this Order. Such drawings, specifications and other documents are to be returned to the Buyer at the completion of the Order or earlier termination of this Agreement. All drawings, specifications and other documents prepared by or for Seller in contemplation of, in the course of, or as a result of performing the work shall be deemed works for hire and all right, title and interest therein shall vest in Buyer, whether or not the Order is ultimately completed. To the extent such drawings, specifications or other documents cannot be considered, by operation of law, works for hire, Seller shall assign to Buyer all right, title and interest thereto and all copies of such drawings, specifications and other documents shall be delivered to Buyer upon completion of the Order or earlier termination of this Agreement. Seller agrees to provide Buyer with reasonable assistance necessary to perfect Seller's interest in intellectual property created under this Agreement. This shall include, but not be limited to, the execution of documents necessary for the Copyright registration. No drawings, specifications or other documents may be used by the Seller or any Sub seller or material or equipment supplier on other projects or for additions to their Project outside the scope of the work without the specific written consent of the Buyer. The Seller, Sub suppliers, Sub-Sub suppliers and material or equipment suppliers are authorized to use and reproduce applicable portions of the drawings, specifications or other documents appropriate to and for use in the execution of their work under the contract documents. All copies made under this authorization shall bear the statutory copyright notice, if any, shown on the drawings, specifications and other documents prepared by or for the Buyer. Submittal or distribution to meet official regulatory requirements or for other purposes in connection with this Project is not to be construed as publication in derogation of the Purchaser's copyrights or other reserved rights. Any intellectual property conceived or developed during the course of the Order based upon or arising from Buyer's confidential and proprietary information shall be solely owned by Buyer. Except as expressly provided herein, no license or right is granted hereby to the Seller, by implication or otherwise, with respect to or under any patent application, patent, claims or patent or proprietary rights of Buyer.</p>
<p>24. The terms and conditions of this Order constitute the entire Agreement between the parties here to and changes will be binding only if the amendments are made in writing and signed by the authorized representatives of the Buyer and the Seller.</p>
<p>25. Risk of loss and/or damage to any goods furnished hereunder shall be upon Seller until the goods are physically delivered to Buyer's facility specified on the face of the Order and accepted by the Buyer.</p>
<p>26. Indemnification: Seller agrees to defend, indemnify and hold harmless the Buyer, its affiliated companies or parent companies, and their officers, employees, agents, guests, invitees and customers from and against any and all liability, loss, damage, fine, penalty, cost or expense (including attorneys' fees) by reason of any allegation, claim, action or suit, whether for death, personal injury, property damage or otherwise, arising out of (1) failure of the goods or services supplied to meet specifications or warranties or for the goods or services to be otherwise defective; or (2) any alleged or actual, direct or contributory infringement or misappropriation of any patent, copyright, trade secret or other proprietary right arising from the purchase, use or sale of such goods or services; or (3) any leak or spill of any goods while being transported or delivered to Buyer; or (4) any breach by Seller of any term or condition contained in the Order; or (5) violation of applicable laws; or (6) alleged defect in the Goods and/or packaging material, or packed Product, or due to the Goods or packaging thereof being alleged to not adhere to any standard or quality set out herein or under any applicable laws; and/or (7) the acts, omissions, or wilful misconduct of Seller's employees and subcontractors, including their agents and representatives, and all other persons performing any services under the Order with the Seller, whether or not caused in part by a party indemnified hereunder. In the event that the goods or services, in Purchaser's reasonable opinion, are likely to infringe a patent or copyright, or misappropriate a trade secret (and in any event, if a court of law finds that the goods or services, in fact, do infringe or misappropriate), then Seller shall further provide Buyer one of the following forms of relief to be chosen by Seller: (a) obtain a license on Buyer's behalf to continue to use or sell the goods or services; (b) redesign the goods or services so that they do</p>

not infringe or misappropriate; or (c) refund Buyer the price paid for the goods or services in question. In any and all claims against Buyer by any employee of Seller, any subcontractor, anyone directly or indirectly employed by any of them, or anyone for whose acts any of them may be liable, the indemnification obligation under the Paragraph shall not be limited in any way by any indemnity or limitation on the amount or type of damages, compensation or benefits payable by or for Supplier, any subcontractor, or anyone directly or indirectly employed by any of them under workers' compensation acts, disability benefit acts, or other employee benefit acts.

27. Confidentiality: Seller shall keep confidential all specifications and proprietary information furnished by Buyer or prepared by Seller in connection with the performance of the Order (including the existence and terms of the Order) and shall not divulge or use such specifications or information for the benefit of itself or any other party, except as required for the efficient performance of the Order. Upon completion of the Order, Supplier shall make no further use, either directly or indirectly, of any such specifications or information.

28. Disposal: If applicable, Seller shall at all times retain title of ownership to any and all materials, substances or chemicals not incorporated into the work that Seller or any subcontractor brings onto Buyer's premises. Seller shall be solely responsible for the handling, transportation and disposal of any and all materials, substances and chemicals. Seller or any subcontractor brings onto Buyer's premises, and any waste generated or resulting from the use thereof. Seller shall not dispose or permit the release of any materials, substance or chemical, or any waste generated or resulting from the use thereof on Buyer's premises. Seller shall handle, transport, and dispose of any and all substances and chemicals, including but not limited to hazardous wastes and substances as defined by applicable federal, state and local laws, rules, regulations, codes and ordinances.

29. Severability: If any provision of this Agreement is held to be invalid, illegal or un-enforceable, either in whole or in part, that holding will not affect the validity, legality or enforceability of the remaining provisions of this Order

30. Original Excise Gate pass must accompany each delivery for excisable goods, if applicable.

31. The Seller will not claim without our knowledge any refund from the excise authorities for the amount of Central Excise duty on the supplies made to us. The Seller shall also undertake to refund to the Buyer all money recovered by him from Govt. authorities for which he has been paid by the Buyer.

32. Unless a specific objection to each of the terms of this Purchase order is raised within 24 hours from the date of Purchase order/email under which this PO is sent, it shall be deemed to be accepted in full.

33. Supplier (Seller) Code of Integrity: The Seller/ Supplier agrees to follow code of integrity and code of conduct as prescribed by General Financial Rules 2017.

TENDER SPECIFICATIONS

S. No	TECHNICAL SPECIFICATIONS	SPECIFICATION AVAILABLE	COMPLIANCE (YES/NO)	THE REASONS/EXPLANATIONS/CONTEXT FOR DEVIATIONS, IF ANY	ADDITIONAL REMARKS FROM THE OEM. YOU CAN USE THIS OPPORTUNITY TO HIGHLIGHT TECHNICAL FEATURES AND QUALIFY THE RESPONSE OF PREVIOUS COLUMNS.	DATASHEET & TECHNICAL OFFER PAGE REFERENCE NUMBER.
A	A latest and most advanced ring shape CZT based SPECT/CT system to be offered by the company.					
1	Scope of Work: The primary vendor shall be responsible for the supply, installation, and commissioning of the SPECT-CT					
2	All necessary support for Quality Assurance (QA) and Acceptance Tests as per NEMA and AERB requirements shall be provided by the company's authorized engineer. The vendor shall assist in preparing and submitting all relevant documentation and test reports to the AERB to obtain the license/approval for operation of the equipment.					
3	The system should be US-FDA/ CE-European with 4 digit notified body number / BIS certified.					
4	The bidder is advised to visit the site and verify that the proposed area allocated for the installation of the equipment is suitable and adequate to meet the equipment's installation and operational requirements.					
4.1	A confirmation email should be attached to say that the space is adequate for the installation of the quoted systems.					
5	The system shall support a comprehensive range of imaging applications, including Dynamic Imaging, SPECT Imaging, Gated Cardiac SPECT, Whole-Body SPECT, and Derived Whole-Body Planar Imaging from SPECT.					
B	General					
1	The system shall be capable of performing Dynamic, SPECT, Gated Cardiac SPECT, Whole-Body SPECT imaging applications.					

2	The system shall be capable of performing high-resolution CT-based attenuation correction and co-registered functional-anatomical image fusion for nuclear medicine data acquisition.					
3	The SPECT/CT system shall have an integrated CT hardware option for CT based attenuation correction and co-registered functional-anatomical image fusion for the same imaging session.					
4	The system shall allow daily and monthly quality control procedures to be performed semi automatically, may require manual intervention for initiation as applicable					
C	Gantry					
1	The integrated SPECT system shall have a gantry bore size of at least 70 cm to accommodate large patients comfortably during imaging					
2	The CT component of the system shall provide a minimum scan field of view (FOV) of 50 cm diameter to ensure full anatomical coverage for large patient sizes.					
3	The gantry should have emergency stop buttons.					
4	Gantry motion should be controlled by a gantry display control panel and user-defined programs.					
5	Height, width, and depth of the system shall be adequate to conveniently locate the gantry within existing departmental space. Manufacturers must provide detailed dimensions for validation.					
6	The system shall be provided with Gantry display & control panel for gantry and detector motion control.					
7	The gantry shall be equipped with safety features including emergency stop buttons on both sides of the gantry					
8	The gantry shall be linked to the patient table with sensors to continuously recognize the patient table position and prevent accidental collisions.					
D	Detectors					
1	The system shall utilize a ring of 10 or more CZT based detectors arranged in a 3D configuration, providing a detector coverage (longitudinal direction) of approximately 26 or more.					

2	The crystal thickness of each detector shall be 7 mm or more					
3	The System shall be equipped with infrared based or equivalent automatic body contouring capability to optimize acquisition geometry.					
4	The system shall provide automatic correction features for energy, and uniformity to maintain image quality.					
5	Performance parameters for the detectors shall comply with NEMA NU 1 latest standards applicable for the model, adhering to the latest specifications prevailing at the time of supply, with supporting literature clearly mentioned.					
8	The CZT energy range of the system shall be specified by the vendor. (At least 250 keV or more)					
11	The system shall be upgradable to incorporate dual-sided CZT (Cadmium Zinc Telluride) detectors capable of performing Iodine based studies.					
14	The intrinsic spatial resolution \leq 2.5mm					
17	The intrinsic energy resolution shall be \leq 6% for Tc ^m 99 & \leq 10% for Lu 177 . Please specify					
20	The intrinsic maximum count rate shall be \geq 400 kcps @ 20% window. Vendor shall specify the exact value.					
23	The reconstructed spatial resolution with scatter (NEMA Nu 1-2018)) in the center direction shall be \leq 5 mm. Vendor shall specify the exact value.					
26	The reconstructed spatial resolution with scatter in the radial direction shall be \leq 5 mm. Vendor shall specify the exact value.					
29	The reconstructed spatial resolution with scatter in the tangential direction shall be \leq 5 mm. Vendor shall specify the exact value.					
E	CT Specifications					
1	The system shall be equipped with a fast-rotation CT scanner, capable of					

	acquiring a minimum of 16 slices per rotation with at least 16 detector rows.					
2	The CT scanner shall generate an attenuation correction map to be applied to nuclear medicine images for attenuation correction and accurate localization by fusing the SPECT and CT images.					
E	The system shall include the latest ultra-low dose technology for attenuation correction, specifically for follow-up and cardiac patients.					
3	Patient bore aperture shall be 70cm.					
4	The scan field of view shall be ≥ 50 cm.					
E	The CT system shall have a continuous X-ray generator supporting sustained and continuous generation, with a capacity of 50kW or more.					
5	Tube current range shall be 15–400 mA or better; vendor to specify exact capability.					
6	The tube voltage range shall be 80–140 kV.					
E	The anode heat storage capacity shall be 6.0 MHU or more.					
7	The scan time for a complete 360-degree rotation shall be 0.5 second or less.					
8	The pitch factor (volume pitch) shall be variable between 0.5 to 1.5 or more, and shall be user selectable.					
9	The system shall allow user selection of combinations of pitch, slice thickness, and spiral length.					
10	The system shall automatically optimize radiation dose and image resolution for each selection.					
11	The single continuous spiral scan time shall be more than 90 seconds.					
12	The slice thickness of the CT reconstructed image shall be ≤ 0.7 mm.					
13	The system shall have 3D mA modulation technology for reducing CT dose.					
14	The system shall offer a metal artifact reduction technique for CT, and the corrected images shall be available for attenuation correction of SPECT data.					
15	The CT subsystem shall be usable for					

	diagnostic CT without requiring an accompanying SPECT study.					
16	The low contrast resolution shall be \leq 3 mm at 3 HU (0.3%) using a 20 cm CATPHAN phantom.					
17	The system shall include a laser alignment light to control the iso-centric position of the patient in all planes.					
F	Collimators					
1	The system shall be provided with Dedicated collimator that is adapted and optimized for CZT detector performance characteristics					
G	Patient Table					
1	The system shall have a single universal patient table suitable for all study types: SPECT, whole-body imaging, and CT images.					
2	The tabletop shall be made of low-attenuation material, preferably carbon fiber. It shall be covered with a mattress pad and straps for patient comfort and safety.					
3	The table shall support a minimum body weight capacity of 200 kg.					
4	The whole-body imaging coverage on the table shall be not less than 180 cm.					
5	Automated patient positioning shall be offered					
6	The table shall move automatically to the home position.					
7	The table height shall be adjustable for easy patient transfer.					
8	The lower vertical motion height of the table shall be 24 inches or less; vendor shall specify the exact value.					
H	Acquisition System					
1	The system shall include one acquisition station capable of acquiring dynamic, multi-gated, whole-body SPECT, and gated SPECT data.					
2	The acquisition station shall consist of a high-performance PC with the latest specifications.					
3	The PC shall run a multi-tasking operating system; vendor shall offer the latest compatible OS version.					
4	The PC shall include an integrated graphics card; vendor					
5	The system shall provide a high-					

	resolution monitor of 1024 × 1024 pixels or higher, sized 17 inches or more, with antiglare flat-panel square LED type; vendor shall specify.					
6	The system shall have a fully integrated CT system capable of acquiring X-ray transmission data together with nuclear emission data, with SPECT and CT data acquisition occurring on the same console.					
7	The system shall be capable of true volumetric SPECT data acquisition without traditional rotating detector heads or step-and-shoot motion.					
8	Acquisition termination shall be controlled by preset time with manual pause, resume, and stop capabilities for all acquisition types.					
9	The system shall have predefined acquisition protocols, along with user-configurable options.					
15	The system shall include a compatible ECG gating device with all leads and cables for gated data acquisition. It shall display ECG and R-R interval histograms during acquisition, indicating frames per R-R interval and maximum frame rate capability.					
16	PHA display					
17	Acquisition software shall include camera quality control (QC) activities:					
17.1	Uniformity correction maps					
17.2	Energy, sensitivity					
17.3	Daily QC protocols					
17.4	Gantry calibration					
18	The acquisition console shall support universal DICOM networking (LAN/WAN) and be able to connect to a network laser color printer.					
19	The system shall provide real-time reconstruction and immediate image display in parallel to acquisition, including in spiral mode.					
I	Server-Client Solution					
1	An OEM Server client solution capable of having at least 3 concurrent users able to run all post processing simultaneously					
2	Server Hardware – Minimum Requirements					
2.1	Processor with minimum 10 cores or higher performance class					

2.2	Minimum 3 TB RAID-enabled storage for data redundancy and reliability					
2.3	Fully redundant, hot-swappable power supplies					
2.4	Fully redundant cooling fans / modules					
2.5	At least two (2) × 1 Gbps Ethernet network ports for connectivity					
2.6	Dedicated graphics card wherever applicable for clinical applications					
2.7	One monitor for service and maintenance purposes					
3	Client PC Hardware – Minimum Requirements					
3.1	Three (3) client systems					
3.2	Minimum 16 GB DDR4 RAM or better (each system)					
3.3	Minimum 1 TB or higher M.2 SSD (each system)					
3.4	Dedicated graphics card as required for clinical and post-processing applications					
3.5	Keyboard and mouse for each client					
3.6	Preloaded licensed operating system					
3.7	All necessary and applicable software licenses for complete functionality					
3.8	Display should be at least 24-inch Dual monitor (2MP Medical grade)					
J	Clinical Applications Software					
1	The system shall support SPECT, whole-body SPECT, including display analysis, Volume rendering with maximum intensity projection (MIP), cine review, curve generation, and manipulation tools.					
2	The system shall include iterative reconstruction, and 3D-OSEM algorithms specifically for SPECT.					
7	The system shall support image output formats of JPEG, TIFF, and AVI, including multimedia reporting and the capability to create self-executable CDs.					
8	The system shall enable annotation, labeling, measurements, and region of interest (ROI) analysis.					
9	The system shall facilitate dynamic evaluation of Tracer enhancement, time-activity curves, time-to-peak images, subtraction, and related techniques.					
12	The system shall provide whole-body SPECT processing software.					

13	The system shall perform complete renal processing including transplant evaluation, diuretic renography, renal split function and NORA.					
14	The system shall provide thyroid volume estimation functionality.					
15	The system shall support Technetium-MIBI subtraction for parathyroid scintigraphy.					
16	The system shall provide condensed dynamic SPECT imaging for and gastric emptying studies.					
17	The system shall include lung perfusion and ventilation imaging VQ SPECT analysis 3D.					
18	The system shall perform bone imaging including three-phase, and SPECT with 3D reconstruction.					
19	The system shall include a full cardiac package supporting first-pass ejection fraction (EF), shunt quantification, gated equilibrium, Gated SPECT, myocardial perfusion, and gated tomography.					
20	The system shall include dedicated licensed cardiac software or toolboxes such as QGS, QPS, QBS/EC Toolbox, or 4DM SPECT for gated cardiac SPECT quantification					
21	The system shall support multi-modality fusion (SPECT/CT/MRI) for anatomical-functional co-registration, including the import of external data.					
22	The system shall support brain processing (3D SPECT)					
23	The system shall include advanced licensed neuroimaging software featuring:					
23.1	MR Fusion					
23.2	Ictal-Interictal subtraction					
24	The system shall support standardized uptake value (SUV) quantification with organ segmentation and ROI transfer between CT and SPECT, based on patient and tracer data.					
25	The system shall include an FDA-approved Voxel based dosimetry application for radiopharmaceutical distribution analysis over time, supporting therapy planning.					
26	Please provide the list of isotopes and their clinical application with white papers					

F	Accessories and QC Utility					
1	System shall be supplied with all required accessories and consumables					
G	ADDITIONAL REQUIREMENTS FOR ALL EQUIPMENT					
1	The procurement and supply of equipment shall be executed in a phased manner based on Client-defined requirements and priorities. The sequence, timelines, and quantity for each phase will be communicated in writing during order award. The Vendor shall comply and ensure timely supply, installation, and commissioning as per the approved phased plan.					
2	The vendor shall provide details of offered AI features					
3	The bidder shall provide a Rate Contract for a minimum of 3 years from the date of supply/installation/commissioning, covering system-specific consumables and accessories.					
4	A complete itemized list of consumables and accessories, including model/reference numbers and units of measurement, shall be submitted in the Technical Offer without prices.					
5	Corresponding unit prices for all consumables and accessories shall be submitted only in the Commercial Offer.					
6	The vendor shall specify the country of origin for the quoted model.					
7	The successful bidder shall submit detailed architectural layouts along with complete MEP (Mechanical, Electrical, and Plumbing) drawings aligned with site conditions and approved drawings. All documents shall be submitted in both editable AutoCAD and PDF formats for review and approval.					

Note: Where datasheet references are not available for any quoted parameter, the bidder shall submit appropriate supporting documents, including product brochures, user manuals, service manuals, technical offers, or a formal declaration confirming that all stated features and functionalities will be demonstrable during Installation, Testing, and Handover (ITC) of the system. All submitted documents shall be clearly identified, cross-referenced, and traceable, with explicit mention of document name, page number, and clause/reference for verification. The supplier shall demonstrate full compliance with all technical specifications, performance parameters, features, and capabilities stated, quoted, claimed, or implied in the bid—whether in compliance statements, technical documents, brochures, datasheets, or any other submitted material—during ITC. Any failure or partial inability to demonstrate any quoted specification shall be treated as technical non-compliance, and the equipment may be rejected and removed from the site.

ANNEXURE:1 ADDITIONAL REQUIREMENTS FOR ALL EQUIPMENT	
1	The procurement and supply of equipment shall be executed in a phased manner, subject to the requirements and priorities as determined by the Client. The sequence, timelines, and quantum of each procurement phase will be communicated in writing by the Client during the awarding of order. The Vendor shall comply with such directives and ensure timely readiness to supply, install, and commission equipment as per the approved phased plan.
2	The vendor should supply middleware for integrating medical equipment with the hospital EMR for interoperability.
3	The vendor shall list the availability of AI features to enhance workflows for all medical equipment as applicable.
4	The bidder shall provide a Rate Contract for 3 years from the date of supply / installation / commissioning, covering system specific consumables and accessories
5	A complete itemized list of all consumables and accessories, including model/reference numbers and unit of measurement, shall be submitted in the Technical Offer (without prices)
6	The corresponding unit prices for the same items shall be submitted only in the Commercial Offer.
7	The vendor should specify the country of origin for the quoted model.

ANNEXURE: 2 SCOPE OF WORK MATRIX					
SN	DESCRIPTION	RESPONSIBILITY OF EQUIPMENT SUPPLIER	RESPONSIBILITY OF CLIENT	ADDITIONAL SCOPE REMARKS	STANDARD /OPTIONAL FOR SUPPLIER
1. DESIGN AND PLANNING					
1.1	SITE SPECIFIC DRAWING WITH PRE-INSTALLATION DETAILS	YES			STANDARD
1.2	APPROVAL OF DRAWINGS AND INCORPORATION INTO THE BUILDING PLAN		YES	NA	
2. CIVIL WORKS AND SITE PREPARATION					
2.1	ENSURE LEVELED FLOOR IN EXAMINATION ROOM AS PER PRE-INSTALLATION REQUIREMENT		YES	NA	
2.2	MAINTAIN REQUIRED FLOOR LEVEL BEFORE INSTALLATION		YES	SUPPLIER SCOPE- ANY MODIFICATION OR FINISHING WORK TO BE CARRIED OUT AFTER INSTALLATION OF EQUIPMENT TO BE QUOTED AS OPTIONAL AS SEPARATE	OPTIONAL COST TO BE QUOTED FOR " ADDITIONAL SCOPE REMARKS"
2.3	WALL CONSTRUCTION AS PER AERB		YES	NA	

	GUIDELINES				
2.4	CONSTRUCTION OF CONCRETE BED IN EXAMINATION ROOM AS PER EQUIPMENT REQUIREMENT		YES	NA	
2.5	FLOOR FINISHING WITH APPROPRIATE MATERIALS		YES	SUPPLIER SCOPE- ANY MODIFICATION OR FINISHING WORK TO BE CARRIED OUT AFTER INSTALLATION OF EQUIPMENT TO BE QUOTED AS OPTIONAL AS SEPARATE	OPTIONAL COST TO BE QUOTED FOR " ADDITIONAL SCOPE REMARKS"
2.6	CUPBOARDS/RACKS WITH LOCKING FACILITY FOR ACCESSORIES, MANUALS, AND TOOLS		YES	NA	
2.7	PAINTING, TILE WORK, AND GENERAL WALL & INTERIOR FINISHING		YES	SUPPLIER SCOPE- ANY MODIFICATION OR FINISHING WORK TO BE CARRIED OUT AFTER INSTALLATION OF EQUIPMENT TO BE QUOTED AS OPTIONAL AS SEPARATE	OPTIONAL COST TO BE QUOTED FOR " ADDITIONAL SCOPE REMARKS"
2.8	CABLE TRAY PROVISION INSIDE FACILITY AS REQUIRED		YES	NA	OPTIONAL COST TO BE QUOTED FOR " 24 ADDITIONAL SCOPE REMARKS"
2.9	RE-CLOSING OF WALLS IF OPENED FOR MOVING IN EQUIPMENT		YES	NA	
3. TRANSPORTATION AND EQUIPMENT MOVEMENT					
3.1	DEFINE AND PREPARE EQUIPMENT TRANSPORTATION ROUTE		YES	NA	
3.2	PROVISION OF CLEAN EQUIPMENT PATH FOR MOBILIZATION UNTIL THE DESIGNATED EQUIPMENT ROOM		YES	NA	
3.3	PROVIDE TRANSPORTATION DIMENSIONS	YES			STANDARD
3.4	PLATFORM FOR UNLOADING AS PER PRE-INSTALLATION REQUIREMENT		YES	NA	
3.5	CONFIRM STRUCTURAL STRENGTH OF UNLOADING PLATFORM AND RIGGING ROUTE		YES	NA	
3.6	ARRIVAL OF EQUIPMENT AT SITE	YES			STANDARD
3.7	ARRANGEMENT OF CRANE FOR UNLOADING	YES			STANDARD

	AS REQUIRED				
3.8	UNLOADING OF EQUIPMENT	YES			STANDARD
3.9	ARRIVAL OF SPECIAL TOOLS AT SITE FOR UNLOADING	YES			STANDARD
4. ELECTRICAL AND POWER REQUIREMENTS					
4.1	POWER SUPPLY OF 415V, 3 PHASE, 50 HZ AS PER PRE-INSTALLATION REQUIREMENT		YES	NA	
4.2	MAINS DISTRIBUTION PANEL TO BE INSTALLED AS PER LINE DIAGRAM		YES	NA	
4.3	TERMINATION OF MAINS CABLE AT MAINS SWITCH (PANEL) IN EQUIPMENT ROOM/UPS ROOM		YES	NA	
4.4	EARTHING AS PER EQUIPMENT PRE-INSTALLATION REQUIREMENT		YES	NA	
4.5	DEDICATE EARTH PIT TO BE PROVIDED AND EARTHING CONNECTION TO BE ESTABLISHED AS PER VENDOR DEFINED LOCATIONS		YES	NA	
9. POWER PANEL FOR EQUIPMENT					
9.1	ALL ELECTRICAL WORK INCLUDING SUPPLY & INSTALLATION OF DISTRIBUTION PANEL FROM MAINS CABLE.		YES	NA	OPTIONAL COST TO BE QUOTED
9.2	SUPPLY OF UPS WITH BATTERIES AND REQUIRED ACCESSORIES FOR THE FUNCTIONING OF SYSTEM	YES			STANDARD
6. LIGHTING AND FALSE CEILING					
6.1	LIGHTING IN ALL ROOMS		YES	NA	
6.2	RADIATION ON LIGHT CONNECTED TO EQUIPMENT AND ITS ASSOCIATED ELECTRICAL WORK AND INSTALLATION	YES			STANDARD
6.3	FALSE CEILING AND INTERIOR FINISHING OF EXAM ROOM		YES	NA	
7. HVAC (AIR CONDITIONING)					
7.1	DUCTING AND TERMINATION		YES	NA	
7.2	MAINTAIN TEMPERATURE AS PER PRE-INSTALLATION REQUIREMENT		YES	NA	
7.3	MAINTAIN RELATIVE HUMIDITY AS PER PRE-INSTALLATION REQUIREMENT		YES	NA	

7.4	AIR-CONDITIONING TO BE FUNCTIONAL AND ENSURE DUST-FREE DIGITAL ENVIRONMENT		YES	NA	
7.5	PROVIDE THERMOMETER WITH TEMPERATURE AND HUMIDITY DISPLAY IN EXAM AND CONSOLE ROOMS	YES			STANDARD
7.6	AIR-CONDITIONING TO BE FUNCTIONAL TO ENSURE DUST-FREE DIGITAL ENVIRONMENT		YES	NA	
8. RADIATION SHIELDING					
8.1	SUPPLY OF LEAD GLASS WINDOW AS PER SITE CONDITION (APPROX 1 M X 1 M) TO BE SUPPLIED AS PER SITE CONDITION	YES			STANDARD
8.2	INSTALL LEAD GLASS WINDOW BETWEEN ROOM AND CONTROL ROOM		YES	SUPPLIER SCOPE- ANY MODIFICATION OR FINISHING WORK TO BE CARRIED OUT AFTER INSTALLATION OF EQUIPMENT TO BE QUOTED AS OPTIONAL AS SEPARATE	OPTIONAL COST TO BE QUOTED FOR " ADDITIONAL SCOPE REMARKS"
8.3	LEAD SHIELDED DOOR AS PER SITE REQUIREMENT		YES	NA	
9. ROOM FINISHES AND ACCESSORIES					
9.1	PROVIDE ANTI-STATIC FLOORING IN EXAMINATION ROOM AFTER EQUIPMENT INSTALLATION		YES	SUPPLIER SCOPE- ANY MODIFICATION OR FINISHING WORK TO BE CARRIED OUT AFTER INSTALLATION OF EQUIPMENT TO BE QUOTED AS OPTIONAL AS SEPARATE	OPTIONAL COST TO BE QUOTED FOR " ADDITIONAL SCOPE REMARKS"
9.2	PROVIDE ANTI-STATIC FLOORING IN CONSOLE ROOM		YES	SUPPLIER SCOPE- ANY MODIFICATION OR FINISHING WORK TO BE CARRIED OUT AFTER INSTALLATION OF EQUIPMENT TO BE QUOTED AS OPTIONAL AS SEPARATE	OPTIONAL COST TO BE QUOTED FOR " ADDITIONAL SCOPE REMARKS"
9.3	PROVIDE REQUIRED NUMBER OF WALL SOCKETS (5A/15A) NEAR CONSOLE AS PER SERVICE PROVIDER REQUIREMENT	VENDOR HAS TO PROVIDE REQUIREMENT	YES		
9.4	PROVIDE WALL SOCKETS (5A/15A) IN EQUIPMENT ROOM	VENDOR HAS TO PROVIDE REQUIREMENT	YES		
9.5	PROVIDE MGPS OUTLETS	VENDOR HAS TO PROVIDE REQUIREMENT	YES		
9.6	ELV SYSTEMS	VENDOR HAS TO PROVIDE	YES		

		REQUIREMENT			
10.	COMMUNICATION				
10.1	INSTALLATION OF STANDALONE INTERCOM AND COMMUNICATION SYSTEM FOR WORKFLOW	YES			STANDARD
11.	CLEANING AND READINESS				
11.1	CLEANLINESS OF SITE AND ENSURING DUST-FREE ENVIRONMENT		YES	NA	
11.2	24 HR POWER SUPPLY		YES	NA	
12.	OTHER ITEMS		YES		
12.1	FIXED FURNITURE		YES	NA	
12.2	LOOSE FURNITURE		YES	NA	
12.3	FIREFIGHTING AND DETECTION SYSTEMS		YES	NA	
12.4	HVAC		YES	NA	
12.5	MEDICAL FURNITURE (TROLLEYS, STOOLS, ETC.)		YES	NA	
12.6	SIGNAGE (DIRECTIONAL, WARNING, DEPARTMENTAL)		YES	NA	
12.7	WORKFLOW-RELATED ITEMS (ZONING SIGNAGE, VISUAL BOARDS, ETC.)		YES	NA	
12.8	TRENCH WORK AS REQUIRED		YES	SUPPLIER SCOPE- ANY MODIFICATION OR FINISHING WORK TO BE CARRIED OUT AFTER INSTALLATION OF EQUIPMENT TO BE QUOTED AS OPTIONAL AS SEPARATE	OPTIONAL COST TO BE QUOTED FOR " ADDITIONAL SCOPE REMARKS"
13.	INSTALLATION				
13.1	INSTALLATION & COMMISSIONING OF EQUIPMENT	YES			STANDARD
14.	STATUTORY COMPLIANCE				
14.1	FACILITATING AERB, PCPNDT WHEREVER APPLICABLE	YES	YES		STANDARD
15	SUPPLIER SCOPE-ANY MODIFICATION OR FINISHING WORK TO BE CARRIED OUT AFTER INSTALLATION OF EQUIPMENT				

ANNEXURE 3: SCOPE OF SUPPLY (FOR TECHNICAL BID)

ANNEXURE 3: SCOPE OF SUPPLY (FOR TECHNICAL BID)									
EQUIPMENT NAME		SPECT/CT							
VENDOR NAME		SPECT/CT							
MAKE		SPECT/CT							
MODEL NAME		SPECT/CT							
SNO	GROUP	ITEM NAME	QUANTITY	VENDOR CATALOGUE NUMBER	STANDARD/OPTIONAL	REMARKS (VENDOR ARE THEIR ITEMS IN THEIR PROFORMA INVOICE WITH ENCLOSED LINE ITEMS)	QUOTE REFERENCE:(SN)	QUOTE PAGE REFERENCE IF APPLICABLE)	
1	HARDWARE	INTEGRATED SPECT/CT SYSTEM WITH MINIMUM 16-SLICE CT	1		STD				
2	HARDWARE	OEM ACQUISITION SYSTEM	1		STD				
3	HARDWARE	SERVER WITH MINIMUM 3 CLIENT WORKSTATIONS	1 SET		STD				
4	HARDWARE	COLLIMATOR SET	1 SET		STD				
5	SOFTWARE	INTEGRATED ECG ACQUISITION AND GATING SYSTEM	1		STD				
6	SOFTWARE	DICOM 3.0 CONNECTIVITY (STORE, PRINT, WORKLIST, MPPS).	AS REQUIRED FOR ENTIRE SYSTEM		STD				
7	SOFTWARE	EXAM HISTORY RETRIEVAL (PACS CONNECTED).	AS REQUIRED FOR ENTIRE SYSTEM		STD				
8	SOFTWARE	MPPS LICENSE – REQUIRED FOR THE CONNECTIVITY BETWEEN THE MODALITY TO THE PACS	AS REQUIRED FOR ENTIRE SYSTEM		STD				
9	SOFTWARE	MODALITY WORKLIST LICENSE – REQUIRED FOR ONLINE ORDERING / SCHEDULING ETC OF THE MODALITY FROM THE PACS	AS REQUIRED FOR ENTIRE SYSTEM		STD				
10	SOFTWARE	OPERATING SYSTEM WITH LIFETIME LICENSE	AS REQUIRED FOR ENTIRE SYSTEM		STD				
11	SOFTWARE	ANTIVIRUS WITH LIFETIME LICENSE	AS REQUIRED FOR ENTIRE SYSTEM		STD				

51.3	WORKFLOW ITEMS	VIAL SHIELD POT 6MM LEAD TO ACCOMMODATE 10ML VIALS	4				STD		
51.4	WORKFLOW ITEMS	VIAL SHIELD POT 6MM LEAD TO ACCOMMODATE 15/20ML VIALS	1				STD		
51.5	WORKFLOW ITEMS	LONG FORCEPS 18" STD	1				STD		
51.6	WORKFLOW ITEMS	SHORT FORCEPS	1				STD		
51.7	WORKFLOW ITEMS	SINGLE SYRINGE HOLDER: LEAD SHIELDING 6MM THICK FOR HOLDING 2-5CC	6				STD		
51.8	WORKFLOW ITEMS	LEAD SYRINGE SHIELD 5CC(6MM) FOOT OPERATED WASTE BIN: 6MM LEAD THICK	1				STD		
51.9	WORKFLOW ITEMS	FOOT OPERATED RECTANGULAR LEAD WASTE BIN WITH HEIGHT, 62CM CAPACITY: 30 LITERS.	2				STD		
51.10	WORKFLOW ITEMS	L-BENCH	1				STD		
51.11	WORKFLOW ITEMS	DECAY DRUM WITH 6 MM THICK LEAD... UNDER CARRIAGE TROLLEY FOR EASY MOBILITY. CAPACITY: 35 LITERS.	1				STD		
51.12	WORKFLOW ITEMS	SYRINGE CARRIER - 6MM LEAD LINED. OD 230MM(L) X 110MM (W) X 84MM (H). LEAD SHIELDING 6MM.	1				STD		
51.13	WORKFLOW ITEMS	DE-CAPPER	1				STD		
51.14	WORKFLOW ITEMS	DOSE CALIBRATOR TOUCH SCREEN:	1				STD		
51.15	WORKFLOW ITEMS	POCKET DOSIMETER 0.001 MSV TO 999.9 MSV	1				STD		
51.16	WORKFLOW ITEMS	DECONTAMINATION KIT	1				STD		
51.17	WORKFLOW ITEMS	DRY BATH	1				STD		
52	WORKFLOW ITEMS	LOW DOSE THERAPY LEAD BRICKS					STD		
52.1	WORKFLOW ITEMS	50MM THICKNESS DIMENSIONS OF 100 LX100 WX50MM THICK	100				STD		

52. 2	WORKFLOW ITEMS	VIAL SHIELD POT 6MM LEAD TO ACCOMMODATE 10ML VIALS	1				STD		
52. 3	WORKFLOW ITEMS	FORCEPS 18" STD	1				STD		
52. 4	WORKFLOW ITEMS	SINGLE SYRINGE HOLDER: LEAD SHIELDING 6MM THICK FOR HOLDING 10CC.	1				STD		
52. 5	WORKFLOW ITEMS	LEAD SYRINGE SHIELD 5CC(6MM) FOOT OPERATED WASTE BIN: 6MM LEAD THICK	2				STD		
52. 6	WORKFLOW ITEMS	FOOT OPERATED RECTANGULAR LEAD WASTE BIN WITH. HEIGHT, 62CM CAPACITY: 30 LITERS.	2				STD		
52. 7	WORKFLOW ITEMS	L-BENCH	1				STD		
52. 8	WORKFLOW ITEMS	DECAY DRUM WITH 6 MM THICK LEAD... UNDER CARRIAGE TROLLEY FOR EASY MOBILITY. CAPACITY: 35 LITERS.	1				STD		
52. 9	WORKFLOW ITEMS	SYRINGE CARRIER - 6MM LEAD LINED. OD 230MM(L) X 110MM (W) X 84MM (H). LEAD SHIELDING 6MM.	1				STD		
52. 10	WORKFLOW ITEMS	DE-CAPPER	2				STD		
52. 11	WORKFLOW ITEMS	DOSE CALIBRATOR TOUCH SCREEN:	1				STD		
52. 12	WORKFLOW ITEMS	DOSE CALIBRATOR INTERLOCKING LEAD RING	15				STD		
52. 13	WORKFLOW ITEMS	POCKET DOSIMETER 0.001 MSV TO 999.9 MSV	1				STD		
52. 14	WORKFLOW ITEMS	DECONTAMINATION KIT	1				STD		
52. 15	WORKFLOW ITEMS	PIPETTE 1000MICRO LITRE	1				STD		
52. 16	WORKFLOW ITEMS	PIPETTE 100 MICRO LITRE	1				STD		
52. 17	WORKFLOW ITEMS	TEST TUBE HOLDER (MULTIPLE SIZED HOLDER)	1				STD		
53	WORKFLOW ITEMS	IODINE THERAPY					STD		
53. 1	WORKFLOW ITEMS	LEAD BRICKS 50MM THICKNESS DIMENSIONS	100				STD		

ITEMS								
3	WORKFLOW ITEMS	MEDICAL REFRIGERATOR (SINGLE DOOR-150-180 LITRE)	1				OPT	
4	UPGRADE	DUAL-SIDED CZT (CADMIUM ZINC TELLURIDE) DETECTORS	1				OPT	
5		ALL OTHER CAPABILITY AS AVAILABLE TO BE ADDED BELOW	1				OPT	

ANNEXURE 4: SCOPE OF SUPPLY (FOR COMMERCIAL BID)												
SNO	GROUP	ITEM NAME	QUANTITY	VENDOR CATALOG NUMBER	STANDARD/OPTIONAL	REMARKS (VENDOR ARE REQUESTED TO MAP THEIR ITEMS IN THEIR PROFORMA INVOICE WITH ENCLOSED LINE ITEMS)	QUOTE REFERENCE(S)	QUOTE REFERENCE(S) IF APPLICABLE)	UNIT COST	TOTAL COST FOR THE QUANTITY MENTIONED	GST %	TOTAL COST WITH GST
1	HARDWARE	INTEGRATED SPECT/CT SYSTEM WITH MINIMUM 16-SLICE CT	1		STD							
2	HARDWARE	OEM ACQUISITION SYSTEM	1		STD							
3	HARDWARE	SERVER WITH MINIMUM 3 CLIENT WORKSTATIONS	1 SET		STD							
4	HARDWARE	COLLIMATOR SET	1 SET		STD							
5	SOFTWARE	INTEGRATED ECG ACQUISITION AND GATING SYSTEM	1		STD							

6	SOFTWARE	DICOM 3.0 CONNECTIVITY (STORE, PRINT, WORKLIST, MPPS).	AS REQUIR ED FOR ENTIRE SYSTEM	STD						
7	SOFTWARE	EXAM HISTORY RETRIEVAL (PACS CONNECTED).	AS REQUIR ED FOR ENTIRE SYSTEM	STD						
8	SOFTWARE	MPPS LICENSE – REQUIRED FOR THE CONNECTIVITY BETWEEN THE MODALITY TO THE PACS	AS REQUIR ED FOR ENTIRE SYSTEM	STD						
9	SOFTWARE	MODALITY WORKLIST LICENSE – REQUIRED FOR ONLINE ORDERING / SCHEDULING ETC OF THE MODALITY FROM THE PACS	AS REQUIR ED FOR ENTIRE SYSTEM	STD						
10	SOFTWARE	OPERATING SYSTEM WITH LIFETIME LICENSE	AS REQUIR ED FOR ENTIRE SYSTEM	STD						
11	SOFTWARE	ANTIVIRUS WITH LIFETIME LICENSE	AS REQUIR ED FOR ENTIRE SYSTEM	STD						
12	SOFTWARE	GENERAL NUCLEAR MEDICINE SOFTWARE PACKAGE	1	STD						
13	SOFTWARE	CARDIAC SPECT SOFTWARE (QGS, QPS, 4DM/EMORY OR EQUIVALENT INCLUDING FUNCTIONAL & QUANTITATIVE ANALYSIS)	1	STD						
14	SOFTWARE	SPECT SUV QUANTIFICATION	1	STD						

52. 6	WORKFLOW ITEMS	FOOT OPERATED WASTE BIN: 6MM LEAD THICK FOOT OPERATED RECTANGULAR LEAD WASTE BIN WITH. HEIGHT. 62CM CAPACITY: 30 LITERS.	2	STD						
52. 7	WORKFLOW ITEMS	L-BENCH	1	STD						
52. 8	WORKFLOW ITEMS	DECAY DRUM WITH 6 MM THICK LEAD... UNDER CARRIAGE TROLLEY FOR EASY MOBILITY. CAPACITY: 35 LITERS.	1	STD						
52. 9	WORKFLOW ITEMS	SYRINGE CARRIER - 6MM LEAD LINED. OD 230MM(L) X 110MM (W) X 84MM (H). LEAD SHIELDING 6MM.	1	STD						
52. 10	WORKFLOW ITEMS	DE-CAPPER	2	STD						
52. 11	WORKFLOW ITEMS	DOSE CALIBRATOR TOUCH SCREEN:	1	STD						
52. 12	WORKFLOW ITEMS	DOSE CALIBRATOR INTERLOCKING LEAD RING	15	STD						
52. 13	WORKFLOW ITEMS	POCKET DOSIMETER 0.001 MSV TO 999.9 MSV	1	STD						
52. 14	WORKFLOW ITEMS	DECONTAMINATION KIT	1	STD						
52. 15	WORKFLOW ITEMS	PIPETTE 1000MICRO LITRE	1	STD						
52. 16	WORKFLOW ITEMS	PIPETTE 100 MICRO LITRE	1	STD						
52. 17	WORKFLOW ITEMS	TEST TUBE HOLDER (MULTIPLE SIZED HOLDER)	1	STD						
53	WORKFLOW ITEMS	IODINE THERAPY LEAD BRICKS 50MM THICKNESS DIMENSIONS OF 100 LX100 WX50MM THICK	100	STD						

	TELLURIDE) DETECTORS									
5	ALL OTHER CAPABILITY AS AVAILABLE TO BE ADDED BELOW	1	OPT							

Annexure 5: Checklist for Technical Bid to be completed and attached along with the tender submission

CHECKLIST FOR VENDOR BEFORE SENDING THE TECHNICAL BID					
Sl. No.	Checklist parameter	Yes/ No	Tender reference		
1	A covering letter, compliance statement, and all pages of the tender document duly signed and sealed by the authorized signatory, as part of the tender compliance, must be enclosed with the technical bid.				
2	Availability of technical proposal need to be provided with separate sealed envelope, mentioning on its envelope IISc tender reference number (PLEASE DO NOT INCLUDE COMMERCIAL BID IN TECHNICAL ENVELOPE)			Section A - point 1	
3	Availability of technical offer (without cost) with model number and make for the quoted model enclosed in technical bid.			Section A - point 1	
4	Availability of the Declaration of warranty period (as required in tender) for the quoted model to be enclosed on the technical bid.			Section A - point 9	
5	Availability of the technical compliance table with six columns for the quoted model to be enclosed on the technical bid. Please provide both pdf and worksheet like excel format			Section A - point 4	
6	Availability of the technical compliance with datasheet and technical offer page number reference for the quoted model to be enclosed on the technical bid.			Section A - point 4. f	

7	Availability of the quoted model technical advantage over comparable equipment from the competitor to be enclosed on the technical bid.	Section A - point 5
8	Availability of the scope of supply (BOQ) as per tender to be enclosed along with technical bid. Please provide both pdf and worksheet like excel format (Excluding cost)	
9	Availability of brochure and any supporting document to validate technical compliance for the quoted model enclosed in technical bid.	Section B - point 8
10	Availability of the technical datasheet for the quoted model, with the relevant specifications highlighted in reference to the Tender technical requirements, must be enclosed with the technical bid.	Section B - point 8
11	Availability of the regulatory certificate (like CDSCO/CE/FDA/ISO/AERB type approval where applicable) for the quoted model to be enclosed on the technical bid.	Section C-Point 17-i
12	Availability of the manufacturer authorization letter for the quoted model to be enclosed on the technical bid where applicable.	Section B - point 1
13	Availability of the list of installation sites with contact details for the quoted model to be enclosed on the technical bid.	Section B - point 5
14	Availability of the confirmation letter on 10 Years of spares support for the quoted model to be enclosed on the technical bid.	Section C - point 5.1
15	Availability of the Details of local service center with technical manpower for the quoted model to be enclosed on the technical bid.	Section C - point 17. f
16	Availability of the Power supply & environmental requirement details for the quoted model to be enclosed on the technical bid.	Section C - point 13. a
17	Availability of the deviation statement from tender terms (with justification) for the quoted model to be enclosed on the technical bid.	Section C - point 18. b
18	The soft copy of technical bid only in both excel and pdf format to be made available in pen drive for the quoted model and enclosed on the technical bid envelope. The pen drive to be labelled with tender reference number and vendor name	Section C - point 19
19	Any open recall or Field Safety Corrective Action (FSCA) associated with the quoted model shall be fully disclosed by the bidder in the technical bid submission.	Section C-Point 3.12
20	Note: Kindly index your technical bid considering the above-mentioned check sheet (not limited) preferably in spiral bound mentioning page number.	

Annexure 6: Checklist for Commercial Bid to be completed and attached along with the tender submission

Checklist for Commercial Bid		Yes/No	Remarks
------------------------------	--	--------	---------

Sl. No.	General Requirements
1	Commercial offer should be in complete alignment with technical offer as mentioned in point no 3 of technical offer checklist
2	Availability of commercial quote need to be provided with separate sealed envelope, mentioning on its envelope IISc tender reference number
3	The scope of supply (BOQ) with commercial details should be in align with technical offer mentioned in point 8 of technical offer checklist
4	The country of origin is clearly mentioned.
5	Word "quote" should be mentioned in the first page instead of Proposal
6	The quote should be signed and sealed. If a digital signature is used, it is clearly indicated
7	The validity period of the quote is clearly mentioned
8	Commercial Quote to be prepared on letter head of the company and it should include
8.1	· Registered office address and billing address
8.2	· Company GST number should be mentioned on the first page
8.3	· Validity
8.4	· Payment Terms – 70% payment on shipment, 20% payment after Installation & commissioning, and remaining 10 % on user satisfaction.
8.5	· Warranty details
8.6	· HSN code of items: Each item shall be listed with its HSN code along with supporting document/literature clarifying the HSN classification.
9	The total amount to be mentioned as unit price, GST percentage, Total price inclusive of tax, total price for total quantity mentioned in the tender)
10	Breakup of cost to be given as annexure for all the line items and it should include:
10.1	· Equipment cost- with GST
10.2	· Accessories- with GST
10.3	· Consumables- with GST
10.4	· Other Items- with GST
11	(Tax should be clearly mentioned as IGST 18% or With CGST 9% and SGST 9% or as applicable)
12	OEM certificate or Authorized distribution letter to be attached
12.1	Additional documents required:
12.1	List of critical spare parts and their estimated unit price. (Item cost should not exceed 30% of the total equipment value)
12.2	Vendor shall provide a supporting document clearly specifying the AMC and CAMC rates as fixed absolute values per year for each of the nine (9) years after the warranty period.
12.3	Rate Contract for 3 years from the date of supply / installation / commissioning, covering all system-specific consumables and accessories.
12.4	Quotation for the one-time maintenance call cost (On call charges)