

GLOBAL TENDER NOTICE

**Multiplex protein array system for cytokine analysis using inverted
fluorescence microscopy**

TENDER DOCUMENT

IISC/DBG/26-27/AR/01

**Department of Developmental Biology and Genetics
Indian Institute of Science, Bangalore**

Table of Contents

Section	Title	Description
Section 1	Bid Schedule	
Section 2	Eligibility Criteria	As specified by IISc
Section 3	Terms and Conditions	As specified by IISc
Section 4	Specifications	Technical Specifications
Section 5	Technical Bid	Annexure I
Section 6	Commercial Bid	Quotation with Price and Technical Specifications

Section 1 – Bid Schedule

Sl. No.	Item	Details
1	Tender Number	IISC/DBG/26-27/AR/01
2	Tender Date	22 nd September 2026
3	Item Description	Supply and Installation of Multiplex protein array system for cytokine analysis using inverted fluorescence microscopy
4	Tender Type	Two bid system: a. Technical Bid (Part A) b. Commercial Bid (Part B)
5	Place of Tender Submission	Office, Department of Developmental Biology and Genetics Indian Institute of Science, Bangalore – 560012 080-2293-2659
6	Last Date and Time for Tender Submission	14 th October 2026
7	For Further Clarification	Department of Developmental Biology and Genetics Indian Institute of Science, Bangalore – 560012 Email: anu@iisc.ac.in , office.dbg@iisc.ac.in

Section 2 – Eligibility Criteria

Prequalification criteria:

1. The Bidder's firm should have existed for at least 5 years. The bidder should enclose company registration certificate.
2. Only the Original Equipment Manufacturer or their authorized representatives across the globe shall participate in the bid.
3. The Bidder must not be blacklisted/banned/suspended or have a record of any service-related dispute with any organization in India or elsewhere. A declaration to this effect has to be given as per Annexure 3.
4. The bidder should sign and submit the declaration for Acceptance of Terms and Conditions as per Annexure 4.
5. The Bidder should provide a support certification as per Annexure 6.
6. The Bidder should have qualified technical service personnel for the instrument(s) based in India.
7. Foreign currency quotations should be on CIP Bangalore basis. INR quotes should be on a FOR-IISc Bangalore basis.
8. The order will be placed only on the bidder who participated in the bid.

Section 3 – Terms and Conditions

A. Submission of Tender:

1. All documentation in the submission should be in English.
2. The bid should be submitted in the two-cover system, i.e., technical bid and commercial bid separately in sealed covers.
3. a. Technical Bid (Part-A) – The technical proposal should contain a compliance table besides the technical specifications listed below. The compliance table should include all the items and in the same order. The first column should describe compliance in a "Yes" or "No" response. The second column should state the extent of deviation if the answer was No or describe the exact specifications if Yes. The third column should state the reasons for deviation if the answer was No. The fourth column can be used to compare your solution with that of competitors or provide details as requested. Suppliers who include any indication of prices in the technical bid will be automatically disqualified.
4. b. Commercial Bid (Part-B) – Commercial bid indicating item-wise price breakdown for the items mentioned in the technical bid, as per the format provided in the tender, and other commercial terms and conditions. The price should be inclusive of all discounts, installation, and training.
5. c. The technical bid and price bid should each be placed in separate sealed covers, superscripting on both envelopes the tender number and the due date. Both sealed covers are to be placed in a bigger cover which should also be sealed and duly superscripted with the Tender No., Tender Description and Due Date.
6. d. The SEALED COVER should reach the address mentioned above on or before the due date mentioned in the tender notice. In case the due date is a holiday, the tender will be accepted and opened on the next working day. If the quotation cover is not sealed, it will be rejected.
7. GST/other taxes, levies, etc. are to be indicated separately. The BIDDER should mention GST Registration and PAN in the tender document and provide a copy of both documents (Indian Bidders only).
8. If price is not quoted in the Commercial Bid as per the format provided in the tender document, the bid is liable to be rejected.
9. The Institute reserves the right to accept or reject any bid and to annul the bidding process and reject all bids at any time prior to the award of contract, without thereby incurring any liability to the affected bidder or bidders or any obligation to inform the affected bidder or bidders.
10. A technical evaluation by the purchase committee may include a demonstration at the Indian Institute of Science to verify the functionalities and capabilities of the system quoted. The technical/purchase committee reserves the right to decide the timeframe within which the demonstration must be completed and to reject bids based on their technical evaluation.
11. All components and accessories should indicate component-wise and itemized breakup. Provide certificates for the country of origin of manufacturing for each line item.
12. Incomplete bids will be summarily rejected.
13. The decision of the purchase committee will be final.

B. Cancellation of Tender:

Notwithstanding anything specified in this tender document, IISc Bangalore, in its sole discretion, unconditionally and without having to assign any reason, reserves the rights:

- A. To accept OR reject the lowest tender or any other tender or all the tenders.
- B. To accept any tender in full or in part.
- C. To reject any tender offer not confirming to the tender terms.

C. Validity of the Offer:

The offer shall be valid for 120 days from the date of opening of the commercial bid.

D. Evaluation of the Offer:

- 1. The technical bid (Part A) will be opened first and evaluated.
- 2. Bidders meeting the required eligibility criteria as stated in Section 2 shall only be considered for Commercial Bid (Part B) opening. Agencies not furnishing the documentary evidence as required will not be considered.
- 3. Pre-qualification of the bidders shall not imply final acceptance of the Commercial Bid. The decision in regard to acceptance and/or rejection of any offer shall be the sole discretion of IISc Bangalore.
- 4. The award of contract will be subject to acceptance of the terms and conditions stated in this tender.
- 5. Any offer which deviates from the vital conditions of the tender is liable to be rejected:
 - (a) Non-submission of complete offers.
 - (b) Receipt of bids after due date and time.
 - (c) Receipt of bids in open conditions.
- 6. In case any BIDDER is silent on any clauses mentioned in these tender documents, IISc Bangalore shall construe that the BIDDER had accepted the clauses as of the tender.
- 7. No revision in the terms and conditions quoted in the offer will be entertained after the last date and time fixed for receipt of tenders.
- 8. Lowest bid will be calculated based on the total price of all essential items and selected optional items tendered for basic equipment along with accessories selected for installation, operation, pre-processing and post-processing, optional items, recommended spares, warranty, and annual maintenance contract.

E. Pre-requisites:

The bidder will provide the prerequisite installation requirement of the equipment along with the technical bid.

F. Warranty:

The complete system is to be under warranty as listed in the technical specifications below, including free supply of consumables, spare parts, and data analysis software, computer from the date of functional installation. Vendor should include cost of any spares that are expected to be needed during the warranty period, including electronics, subcomponents, and software. If the instrument is found to be defective, it must be replaced or rectified at the cost of the bidder within 30 days from receipt of written communications from IISc Bangalore.

G. Annual Maintenance Contract (AMC):

An annual maintenance contract as listed in the technical specifications below.

H. Purchase Order:

1. The order will be placed on the bidder whose bid is accepted by IISc Bangalore based on the terms and conditions mentioned in the tender document.
2. The quantity of the items in the tender is only indicative. IISc Bangalore reserves the right to increase/decrease the quantity of items depending on the requirement.
3. If the quality of the product and service provided is not found satisfactory, IISc Bangalore reserves the right to cancel or amend the contract.
4. After the award of the purchase order, the vendor must provide an Order Acknowledgement within 30 days from the receipt of the Purchase Order.

I. Delivery, Installation and Training:

The system should be delivered, installed, and made functional within 10–12 weeks from the date of receipt of purchase order. Supply will be considered complete only on satisfactory installation and inspection of the system and all features/capabilities tested by IISc Bangalore. The date of taking over of the entire system shall mark the start of the warranty period. No partial shipment is allowed. The bidder must provide on-site technical training to facility technologists and users at no extra cost.

J. Payment Terms:

For INR quotes, 100% payments will be released after completion of delivery, satisfactory installation, and qualification, subject to TDS as per rules. As per GFR, no advance payment can be made to domestic vendors. For INR quotes, the price must be on FOR IISc Bangalore basis only.

K. Statutory Variation:

Any statutory increase in taxes and duties subsequent to the bidder's offer, if within the original contractual delivery date, will be borne by IISc Bangalore subject to supporting documentary evidence. Any decrease will be passed on to IISc Bangalore.

L. Dispute and Jurisdiction:

Any legal disputes arising out of any breach of contract pertaining to this tender shall be settled in the court of competent jurisdiction located within the city of Bangalore, India.

M. General:

1. All amendments, time extensions, and clarifications within the period of submission will be communicated electronically.
2. The bidder may furnish any additional information necessary to establish capabilities to successfully complete the envisaged work.
3. The bidder may visit the installation site before submission of tender, with prior intimation.
4. Any information furnished by the bidder found to be incorrect would render the bidder liable to be debarred from tendering/taking up work in IISc Bangalore.

Section 4 – Technical Specifications

Multiplex protein array system for cytokine analysis using inverted fluorescence microscopy

Essential Specifications

The system is primarily intended for quantitative multiplex fluorescence-based detection and analysis of cytokines and other protein biomarkers in assay slides, microarrays, multi-well plates, or compatible multiplex assay substrates containing processed human plasma and cerebrospinal fluid samples. The system shall enable automated fluorescence acquisition, multi-region scanning, image stitching, background correction, quantitative intensity measurement, and export of cytokine assay data. The additional transmitted-light, environmental-control, live-sample imaging, and advanced-imaging capabilities specified below are required to support broader institutional use and future research applications.

#	Parameter	Specification
1	Microscope Frame	a) Fully motorized rugged inverted fluorescence microscope body with spill-proof technology for research applications suitable for automated fluorescence imaging and quantitative analysis of multiplex cytokine and protein assay slides, microarrays, chambered substrates, and multi-well plates. b) Motorized ergonomic stand with in-built Z-focus drive with minimum step resolution of 10 nm or better. c) The system should have a dedicated remote pad for controlling all motorized functions of the microscope. d) Infinity-corrected optical system with full expandability for fluorescence, transmitted light, and advanced imaging modalities with a maximum field of view of 25 mm or better. e) System should have active right and left side port with motorized light path switching between eye and camera (Eyepiece 100%, Left 100%, Right 100%, Eyepiece 50%/Left 50%). f) Coded intermediate magnification changer for 3 positions: 1×, 1.6×, 2×. g) Tilting binocular observation tube with 35° to 85° tilting angle, with a pair of 10× eyepieces of Field Number 22 or more, and dioptre adjustment for both eyes. h) C-mount adapter 1×, video port adapter. i) Dust cover.
2	Transmitted Illumination	High-power, long-working-life LED illumination for transmitted light microscopy. Average LED lifetime ≥50,000 hours.
3	Condenser	Motorized 7-position (or more) long-working-distance (LWD) condenser with NA ≥0.55 and working distance ≥27 mm. Motorized polarizer and motorized mechanical shutter in the transmitted light path. Phase rings compatible with 10× and 20× objectives to be supplied.
4	Nosepiece & Objectives	Six-position motorized revolving nosepiece to accommodate objectives of different magnifications. Objectives to be supplied:

		a) Semi-apochromat 10× phase contrast objective, NA ≥0.30, WD ≥10 mm, resolution required : <1.3μm b) Fluorite/semi-apochromat long-working-distance phase contrast 20×/NA 0.45, WD 6.5–7.5 mm, with correction collar for imaging in thick coverslip or plastic-bottom dishes. resolution required : <1μm c) Fluorite/semi-apochromat long-working-distance 40×/NA 0.6, WD 3.0–4.0 mm (or more), with correction collar for imaging in thick coverslip or plastic-bottom dishes. resolution required : <0.5μm d) A 4× or 5× plan fluorite objective, NA>0.1 suitable for low-magnification fluorescence scanning. The bidder shall specify the achievable lateral resolution, effective pixel size, and field of view for the objectives There should be a possibility to upgrade at least one nosepiece position to a motorized/automatic correction collar, along with the corresponding software module, as a future upgrade.
5	Motorized XY Stage	High-precision motorized XY scanning stage for precise multipoint/multi-well imaging and seamless stitching of large specimens (cytokine assay slides, protein microarrays, antibody arrays), with control through joystick and imaging software. Stage inserts to be supplied for: glass slides, 96-well plate, 35 mm dish, 60 mm dish, and chambered slide — for multi-well multipoint imaging and automated stitching in XYZ and multichannel mode. The bidder shall specify the XY travel range, minimum step size, positioning accuracy, bidirectional repeatability, and maximum supported sample weight.
6	Fluorescence Module	Motorized fluorescence cube turret with integrated motorized shutter and a minimum of 7–8 filter cube slots. Filter sets for DAPI, FITC/GFP, and TRITC must be supplied. The system must be on-site upgradable to additional channels (e.g., CFP, YFP, mCherry) without the need to remove or replace existing four filter cubes. At least one empty spare cube slot to be available.
7	LED Fluorescence Light Source	Multi-channel LED fluorescence illumination system with at least 4 independent LED channels for DAPI, GFP/FITC and TRITC/Cy3. Delivered via liquid light guide and collimator. Average LED lifetime ≥25,000 hours. Software-controlled via imaging software.
8	Z-Drift Compensation System	Hardware-based Z-drift compensation system using an IR diode laser (wavelength: 830 nm). Continuous and single-shot operation modes must be supported. Compatible observation methods: fluorescence, DIC, and phase contrast. Must include IR cut filter and weight for stabilising sample vessel.
9	Camera	Back-side-illuminated (BSI) monochrome digital camera, 6.4 MP resolution. Sensor: 1/1.8-inch monochrome CMOS. Pixel size: 2.4 μm × 2.4 μm. Dynamic range: ≥10 bit. Exposure: manual and auto, 13 μs–15 s. Interface: USB 3.1 Type-C. Frame rate: ≥45 fps (full resolution), ≥59 fps (2×2 binning), ≥60 fps (full HD).

		High-resolution acquisition: ≥ 30 fps for fluorescence, phase contrast, and DIC images.
10	Image Acquisition & Analysis Software	The primary software application shall be automated fluorescence acquisition and quantitative analysis of multiplex cytokine and protein assay substrates. The additional multidimensional, time-lapse, perfusion-control, live-sample, and advanced-imaging functions listed below are required for shared institutional use and future expansion. The imaging software must control all motorized functions of the microscope, camera, and fluorescence light source. Modules and capabilities required: a) Drivers/modules for third-party TTL/triggered device control and DAQ card control for external light sources/lasers and perfusion systems. b) On-stage incubator control and compatibility for future hardware upgrades such as TIRF, FRAP, confocal, and super-resolution systems. c) Multidimensional (XYZt) image acquisition. d) Analysis features: intensity measurement, fluorescence unmixing, co-localisation, time-lapse recording, basic measurements (length, area, perimeter), and data export to Excel/Word. e) Intelligent shading correction for high-quality real-time stitched image creation. f) 3D volume rendering and real-time EDF/EFI (extended depth of focus/field) to create 2D projected images from Z-stacks in real time. g) A dedicated Experiment Manager / Experiment Designer / JOBS module (catalogue-specified; equivalent modules are not acceptable) for creating custom experimental workflows using drag-and-drop functions. h) Macro creation and batch macro processing. i) Cropping, background subtraction, and 2D deconvolution. j) User experience customisation, overlay multiple images, document groups for side-by-side image comparison, movie playback, tile image, and orthogonal slice view of 3D/time-lapse data. k) Active photobleaching-prevention mode. l) High Dynamic Range (HDR) imaging and Dynamic ROI for intensity analysis of moving objects. m) Well-plate module for automated multi-well imaging. n) Multi-position module. Software must be compatible with Windows 11 Pro 64-bit.
11	Stage-top environmental-control	Stage-top incubation system for sample maintenance: a) Stage-top chamber providing optimal sample temperature, humidity, and CO₂ or similar environment. b) Controller with built-in gas mixer for 100% CO₂ cylinder use; automatically mixes CO₂ with surrounding air and supplies 5.0–20.0% CO₂ concentration to the chamber or similar. c) Compatible with multi-well plates and small vessels. d) 35 mm and 50/60 mm petri dish sample holders to be included. e) Top-heater heating method for uniform temperature distribution. f) Internal humidifier for humidity $\geq 90\%$. g) Chamber with at least 3 access ports for perfusion/media exchange (IN and OUT), drug administration, and temperature sensing.

		h) Continuous Current Control to minimise Z-drift caused by heating. i) Objective heater included. j) Dish attachment and fixing lid for slide glass, chamber slide, and chambered cover glass.
12	Acquisition Workstation	Dell Precision 5860 Tower Workstation or equivalent, with: • Intel Xeon processor (or equivalent high-end processor) • 32 GB RAM • 2 TB HDD + 512 GB SSD • DVD-RW drive • Gigabit Ethernet • Nvidia GPU ≥ 4 GB VRAM (or equivalent) • USB keyboard and mouse • Windows 11 Pro 64-bit • 27-inch monitor or equivalent
13	Power Source	System must be directly usable with 230 V / 50 Hz mains supply.
14	Certification	Equipment must carry valid ISO/CE certification. A copy of the valid certificate must be provided with the bid.
15	Customer Base	Similar or identical system must have been supplied to IISc in the past 5-10 years and details of the labs to be furnished in addition at least five universities/institutes globally. Names and contact details of at least five referee users at these organisations must be provided.
16	Warranty	Minimum 3 years comprehensive warranty from the date of functional installation, covering all components, electronics, and software. Cost of spares expected during the warranty period must be included. And 2 year AMC at the end of warranty.
17	Support	Training and instructions on operation of the system to be provided by qualified engineers of the bidder's firm to IISc technical staff, free of cost, after installation. Dedicated service and application support personnel must be based in Bangalore. Please provide the number and details of trained personnel available for this instrument.
18	Breakdown Response	Service call within 48 hours of complaint registration.

Optional Accessories

#	Parameter	Specification
19	Motorized Correction Collar Upgrade	Motorized/automatic correction collar for one nosepiece position, with corresponding software module.
20	Additional Objectives	The bidder may also quote suitable low- and intermediate-magnification objectives optimised for fluorescence scanning, large field of view, flat-field performance, and quantitative imaging of protein arrays or multiplex assay substrates. For every optional objective, the bidder shall provide magnification, numerical aperture, working distance, correction range, immersion medium, optical resolution, field of view, effective specimen-plane pixel size, and intended application.

21	Anti-Vibration Table	Passive air-isolation optical table compatible with the microscope footprint, including low-noise air compressor. Size, flatness specifications, and mounting hole pattern to be specified by vendor.
22	Additional Fluorescence Channels	Filter cubes for additional dyes (e.g., CY5, CFP, YFP, mCherry, Cy7) installable without removing existing filter cubes. Please quote individually.
23	Software	3D Blind Deconvolution Electronic license module
24	Immersion oil bottle 30c	Immersion oil bottle 30cc. Low level autofluorescence

Section 5 – Technical Bid

The technical bid should furnish all requirements of the tender along with all annexures in this section and be submitted to:

The Chairperson,
Department of Developmental Biology and Genetics,
Indian Institute of Science,
Bangalore – 560012, India.

Annexure 1 – Details of the Bidder

The bidder must provide the following mandatory information and attach supporting documents wherever mentioned:

Sl. No.	Item	Details
1	Name of the Bidder	
2	Nature of Bidder (Attach attested copy of Certificate of Incorporation/Partnership Deed)	
3	Registration No. / Trade License (attach attested copy)	
4	Registered Office Address	
5	Address for Communication	
6	Email ID	
7	Website	
8	Contact Person – Name and Designation	
9	Telephone No.	
10	PAN No. (attach copy) – For Indian Vendor only	
11	GST No. (attach copy) – For Indian Vendor only	

Signature of the Bidder: _____

Name: _____

Designation, Seal: Date:

Annexure 2 – Declaration Regarding Experience

To,
The Chairperson,
Department of Developmental Biology and Genetics,
Indian Institute of Science,
Bangalore – 560012, India.

Ref: Tender No: XXXXXXXXXX Dated: XXXXX

Dear Sir/Madam,

I have carefully reviewed the Terms & Conditions in the above-referred tender. I hereby declare that my company/firm has _____ years of experience in supplying and installing the proposed equipment.

Yours faithfully,

(Signature of the Bidder)

Name:

Designation, Seal

Date:

Annexure 3 – Declaration Regarding Track Record

To,
The Chairperson,
Department of Developmental Biology and Genetics,
Indian Institute of Science, Bangalore – 560012, India.

Ref: Tender No: XXXXXXXX Dated: XXXXX

Dear Sir/Madam,

I have carefully reviewed the Terms & Conditions in the above-referred tender. I hereby declare that my company/firm is not currently debarred/blacklisted by any Government/Semi-Government organizations/institutions in India or abroad. I further certify that I am a competent officer in my company/firm to make this declaration.

Or

I declare the following:

Sl. No.	Country in which the company is Debarred/Blacklisted/case Pending	Blacklisted/Debarred by Government/Semi-Government/Organizations	Reason	Since when and for how long
1				
2				

(NOTE: In case the company/firm was blacklisted previously, please provide the details regarding the period and reason/s for the same.)

Yours faithfully,

(Signature of the Bidder)

Name:

Designation, Seal:

Date:

Annexure 4 – Declaration for Acceptance of Terms and Conditions

To,
The Chairperson,
Department of Developmental Biology and Genetics,
Indian Institute of Science,
Bangalore – 560012, India.

Ref: Tender No: XXXXXXXXXX Dated: XXXXX

Dear Sir/Madam,

I have carefully reviewed the Terms & Conditions mentioned in the above-referred tender document. I declare that all the provisions of this tender document are acceptable to my company. I further certify that I am an authorized signatory of my company and am, therefore, competent to make this declaration.

Yours faithfully,

(Signature of the Bidder)

Name:

Designation, Seal

Date:

Annexure 5 – Details of Items Quoted

1. Company Name:
2. Product Name:
3. Part / Catalogue Number:
4. Product Description / Main Features:
5. Detailed Technical Specifications:
6. Remarks:

Instructions to Bidders:

1. Bidder should provide technical specifications of the quoted product(s) in detail.
2. Bidder should attach product brochures along with the technical bid.
3. Bidders should clearly indicate compliance or non-compliance with the technical specifications provided in the tender document.

Annexure 6 – Support Certification

To,
The Chairperson,
Department of Developmental Biology and Genetics,
Indian Institute of Science,
Bangalore – 560012, India.

Ref: Tender No: XXXXXXXXXX Dated: XXXXX

Dear Sir/Madam,

I have carefully reviewed the Terms & Conditions mentioned in the above-referred tender document. I declare that the model quoted has not been discontinued by our organisation and the repair parts will be available for at least another ten years. I further certify that I am an authorized signatory of my company and am, therefore, competent to make this declaration.

Yours faithfully,

(Signature of the Bidder)

Name:

Designation, Seal

Date:

Section 6 – Commercial Bid

The commercial bid should be furnished with all requirements of the tender with supporting documents as mentioned under:

S. No.	Description	Cat. Number	Qty	Unit Price	Sub Total
1	Essential items noted in the technical specification				
1a	Motorized inverted microscope frame with spill-proof body, 10 nm Z-focus, remote pad, coded magnification changer (1×/1.6×/2×)		1		
1b	Tilting binocular observation tube (35°–85°), 10× eyepieces FN 22, C-mount adapter, video port adapter, dust cover		1		
1c	Transmitted LED illumination system (long-life, high-power)		1		
1d	Motorized 7-position LWD condenser (NA ≥0.55, WD ≥27 mm) with motorized polarizer and motorized mechanical shutter		1		
1e	Phase contrast ring slits for 10× and 20× objectives		1 set		
1f	10× Semi-apochromat phase objective (NA ≥0.30, WD ≥10 mm)		1		
1g	20× LWD Fluorite/Semi-apochromat phase objective (NA 0.45, WD 6.5–7.5 mm, correction collar)		1		
1h	40× LWD Fluorite/Semi-apochromat objective (NA 0.6, WD 3.0–4.0 mm, correction collar)		1		
1i	Motorized XY scanning stage with joystick control		1		
1j	Stage inserts: glass slide, 96-well plate, 35 mm dish, 60 mm dish, chambered slide		1 set		
1k	7–8 position motorized fluorescence cube turret with integrated motorized shutter		1		
1l	Fluorescence filter sets: DAPI, FITC/GFP, TRITC + empty cube(s)		1 set		
1m	4-channel LED fluorescence light source (DAPI, FITC, TRITC) with liquid light guide		1		
1n	Z-drift compensation system (hardware IR 830 nm laser-based)		1		

1o	BSI monochrome sCMOS camera (6.4 MP, ≥45 fps, USB 3.1)		1		
1p	Image acquisition & analysis software (5D, XYZt, Experiment Manager, HDR, EFI, deconvolution, unmixing, TTL/DAQ control, macro, TIRF/FRAP upgrade-ready, 3D blind deconvolution, well-plate & multi-position modules)		1		
1q	Stage-top incubator with built-in gas mixer, objective heater, dish/plate adapters, humidifier		1		
1r	Acquisition workstation (Intel Xeon/equiv., 32 GB RAM, 2 TB HDD + 512 GB SSD, Nvidia GPU, Win 11 Pro, 27" monitor)		1		
2	Optional items noted in the technical specification				
2a	Annual Maintenance Contract – 1 year post-warranty		1		
2b	Motorized/automatic correction collar upgrade (1 nosepiece position) with software module		1		
2c	Anti-vibration table with low-noise air compressor		1		
2d	Additional fluorescence filter cubes (e.g., CFP, YFP, mCherry, Cy7) – please specify individually				
2e	Additional objectives (please specify)				
3	Accessories for operation and installation				
4	All consumables, spares and software to be supplied locally				
5	Warranty (minimum 1 year)				
6	Cost of insurance and freight (if applicable)				
7	CIP/CIF/FOR IISc, Bengaluru (grand total before GST)				
8	GST (please specify applicable %)				
	TOTAL LANDING COST (inclusive of all taxes)				

Addressed to:

The Chairperson,

Department of Developmental Biology and Genetics,

Indian Institute of Science, Bangalore – 560012, India.

Section 7 – Checklist

(This should be enclosed with the Technical Bid – Part A)

The following items must be checked before the Bid is submitted:

A. Sealed Envelope "A": Technical Bid

1. Section 5 – Technical Bid (each page signed by the authorized signatory and sealed) with the below annexures:
2. Annexure 1: Bidder's details
3. Annexure 2: Declaration regarding experience
4. Annexure 3: Declaration regarding clean track record
5. Annexure 4: Declaration for acceptance of terms and conditions
6. Annexure 5: Details of items quoted
7. Annexure 6: Support certification
8. Copy of this tender document duly signed by the authorized signatory on every page and sealed.

B. Sealed Envelope "B": Commercial Bid

Section 6: Commercial Bid

Your quotation must be submitted in two envelopes: Technical Bid (Envelope A) and Commercial Bid (Envelope B), superscribing on both the envelopes with Tender No. and due date, both in sealed covers placed in a bigger cover which should also be sealed and duly superscribed with Tender No., Tender Description and Due Date.